

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031422\
 Data File : VN071271.D
 Acq On : 14 Mar 2022 17:14
 Operator : JC\MD
 Sample : PB143286ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143286ZHE#03

Quant Time: Mar 15 02:53:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	725087	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1191165	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1101566	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	400464	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	459597	46.248	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.500%
35) Dibromofluoromethane	8.016	113	347240	46.196	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.400%
50) Toluene-d8	10.439	98	1479275	48.878	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.760%
62) 4-Bromofluorobenzene	12.727	95	487331	47.553	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.100%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	34613	7.158	ug/l	98
18) Methyl Acetate	4.851	43	26451	2.488	ug/l	96
20) Methylene Chloride	5.087	84	15410	2.112	ug/l	94
43) Isopropyl Acetate	8.563	43	126439	5.101	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031422\
Data File : VN071271.D
Acq On : 14 Mar 2022 17:14
Operator : JC\MD
Sample : PB143286ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143286ZHE#03

Quant Time: Mar 15 02:53:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

