

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031422\
 Data File : VN071272.D
 Acq On : 14 Mar 2022 17:38
 Operator : JC\MD
 Sample : PB143286ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143286ZHE#04

Manual Integrations
 APPROVED

Reviewed By : John Carlone 03/15/2022
 Supervised By : Mahesh Dadoda 03/15/2022

Quant Time: Mar 15 02:53:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	775837	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1284557	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1159893	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	394061	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	487429	45.841	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	91.680%		
35) Dibromofluoromethane	8.016	113	369120	45.537	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	91.080%		
50) Toluene-d8	10.439	98	1567684	48.034	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.060%		
62) 4-Bromofluorobenzene	12.727	95	490961	44.424	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	88.840%		
Target Compounds					Qvalue	
16) Acetone	4.287	43	36749	7.103	ug/l	95
18) Methyl Acetate	4.857	43	32012	2.752	ug/l	97
20) Methylene Chloride	5.081	84	15710m	2.031	ug/l	
43) Isopropyl Acetate	8.563	43	167223	6.256	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031422\
Data File : VN071272.D
Acq On : 14 Mar 2022 17:38
Operator : JC\MD
Sample : PB143286ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143286ZHE#04

Manual Integrations
APPROVED

Reviewed By :John Carlone 03/15/2022
Supervised By :Mahesh Dadoda 03/15/2022

Quant Time: Mar 15 02:53:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

