

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031422\
 Data File : VN071273.D
 Acq On : 14 Mar 2022 18:02
 Operator : JC\MD
 Sample : PB143286ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143286ZHE#05

Quant Time: Mar 15 02:54:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	708876	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1195946	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1110926	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	397442	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	463853	47.744	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.480%
35) Dibromofluoromethane	8.022	113	349532	46.315	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.640%
50) Toluene-d8	10.439	98	1480815	48.734	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.460%
62) 4-Bromofluorobenzene	12.727	95	477457	46.403	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.800%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	35642	7.540	ug/l	97
18) Methyl Acetate	4.851	43	32784	3.028	ug/l	95
20) Methylene Chloride	5.098	84	15799	2.197	ug/l #	89
43) Isopropyl Acetate	8.569	43	213097	8.563	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031422\
Data File : VN071273.D
Acq On : 14 Mar 2022 18:02
Operator : JC\MD
Sample : PB143286ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143286ZHE#05

Quant Time: Mar 15 02:54:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

