

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031422\
 Data File : VN071274.D
 Acq On : 14 Mar 2022 18:26
 Operator : JC\MD
 Sample : PB143286ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143286ZHE#06

Quant Time: Mar 15 02:54:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	781499	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1309293	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1198596	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	439303	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	500437	46.723	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.440%
35) Dibromofluoromethane	8.016	113	383453	46.411	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.820%
50) Toluene-d8	10.439	98	1614110	48.522	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.040%
62) 4-Bromofluorobenzene	12.727	95	536283	47.608	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.220%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	33022	6.336	ug/l	95
18) Methyl Acetate	4.863	43	35160	2.959	ug/l	93
20) Methylene Chloride	5.092	84	14896	1.934	ug/l #	93
43) Isopropyl Acetate	8.563	43	179987	6.607	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031422\
Data File : VN071274.D
Acq On : 14 Mar 2022 18:26
Operator : JC\MD
Sample : PB143286ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143286ZHE#06

Quant Time: Mar 15 02:54:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

