

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031422\
 Data File : VN071275.D
 Acq On : 14 Mar 2022 18:50
 Operator : JC\MD
 Sample : PB143286ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143286ZHE#07

Quant Time: Mar 15 02:54:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	759862	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1284374	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1192695	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	430808	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	485981	46.665	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.340%
35) Dibromofluoromethane	8.016	113	377223	46.543	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.080%
50) Toluene-d8	10.439	98	1600123	49.035	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.060%
62) 4-Bromofluorobenzene	12.727	95	520510	47.104	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.200%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	33825	6.675	ug/l	98
18) Methyl Acetate	4.857	43	34956	3.015	ug/l	96
20) Methylene Chloride	5.098	84	14935	1.982	ug/l #	83
43) Isopropyl Acetate	8.563	43	277192	10.372	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031422\
Data File : VN071275.D
Acq On : 14 Mar 2022 18:50
Operator : JC\MD
Sample : PB143286ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143286ZHE#07

Quant Time: Mar 15 02:54:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

