

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031422\
 Data File : VN071277.D
 Acq On : 14 Mar 2022 19:39
 Operator : JC\MD
 Sample : PB143286ZHE#09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143286ZHE#09

Quant Time: Mar 15 02:55:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	781598	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1327279	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1216105	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	443423	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	498038	46.493	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	92.980%		
35) Dibromofluoromethane	8.016	113	384864	45.951	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	91.900%		
50) Toluene-d8	10.439	98	1617039	47.951	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	95.900%		
62) 4-Bromofluorobenzene	12.727	95	532851	46.662	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	93.320%		
Target Compounds						
16) Acetone	4.293	43	34881	6.692	ug/l	99
18) Methyl Acetate	4.857	43	37820	3.146	ug/l	94
20) Methylene Chloride	5.098	84	14467	1.889	ug/l	97
43) Isopropyl Acetate	8.569	43	258094	9.345	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031422\
Data File : VN071277.D
Acq On : 14 Mar 2022 19:39
Operator : JC\MD
Sample : PB143286ZHE#09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143286ZHE#09

Quant Time: Mar 15 02:55:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

