

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031422\
 Data File : VN071279.D
 Acq On : 14 Mar 2022 20:27
 Operator : JC\MD
 Sample : PB143286ZHE#11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143286ZHE#11

Quant Time: Mar 15 02:55:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	739209	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1247158	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1124904	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	400393	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	473818	46.769	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.540%
35) Dibromofluoromethane	8.022	113	363493	46.187	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.380%
50) Toluene-d8	10.439	98	1534662	48.432	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.860%
62) 4-Bromofluorobenzene	12.727	95	484079	45.115	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.220%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	35008	7.102	ug/l #	85
18) Methyl Acetate	4.857	43	40146	3.474	ug/l	93
20) Methylene Chloride	5.098	84	14893	2.022	ug/l #	91
43) Isopropyl Acetate	8.563	43	286698	11.048	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031422\
Data File : VN071279.D
Acq On : 14 Mar 2022 20:27
Operator : JC\MD
Sample : PB143286ZHE#11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143286ZHE#11

Quant Time: Mar 15 02:55:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

