

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031422\
 Data File : VN071282.D
 Acq On : 14 Mar 2022 21:39
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 15 04:03:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	0.548	0.494	9.9	83	0.00
3 P	Chloromethane	0.822	0.718	12.7	86	0.00
4 C	Vinyl Chloride	0.672	0.660	1.8#	90	0.00
5 T	Bromomethane	0.335	0.350	-4.5	90	0.00
6 T	Chloroethane	0.407	0.403	1.0	86	0.00
7 T	Trichlorofluoromethane	0.884	0.861	2.6	89	0.00
8 T	Diethyl Ether	0.378	0.383	-1.3	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.512	0.496	3.1	89	0.00
10 T	Methyl Iodide	0.723	0.688	4.8	83	0.00
11 T	Tert butyl alcohol	0.135	0.144	-6.7	98	0.00
12 CM	1,1-Dichloroethene	0.501	0.492	1.8#	89	0.00
13 T	Acrolein	0.142	0.142	0.0	96	0.00
14 T	Allyl chloride	1.090	1.073	1.6	89	0.00
15 T	Acrylonitrile	0.400	0.407	-1.7	91	0.00
16 T	Acetone	0.333	0.347	-4.2	96	0.00
17 T	Carbon Disulfide	1.427	1.256	12.0	80	0.00
18 T	Methyl Acetate	0.979	0.999	-2.0	99	0.00
19 T	Methyl tert-butyl Ether	1.882	1.988	-5.6	93	0.00
20 T	Methylene Chloride	0.661	0.614	7.1	89	0.00
21 T	trans-1,2-Dichloroethene	0.548	0.534	2.6	88	0.00
22 T	Diisopropyl ether	2.148	2.197	-2.3	91	0.00
23 T	Vinyl Acetate	1.818	1.926	-5.9	91	0.00
24 P	1,1-Dichloroethane	1.135	1.149	-1.2	92	0.00
25 T	2-Butanone	0.537	0.569	-6.0	94	0.00
26 T	2,2-Dichloropropane	0.876	0.828	5.5	84	0.00
27 T	cis-1,2-Dichloroethene	0.651	0.656	-0.8	91	0.00
28 T	Bromochloromethane	0.505	0.476	5.7	84	0.00
29 T	Tetrahydrofuran	0.372	0.386	-3.8	92	0.00
30 C	Chloroform	1.083	1.097	-1.3#	92	0.00
31 T	Cyclohexane	1.094	1.024	6.4	87	0.00
32 T	1,1,1-Trichloroethane	0.922	0.955	-3.6	92	0.00
33 S	1,2-Dichloroethane-d4	0.685	0.658	3.9	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.316	0.304	3.8	91	0.00
36 T	1,1-Dichloropropene	0.506	0.490	3.2	89	0.00
37 T	Ethyl Acetate	0.696	0.673	3.3	87	0.00
38 T	Carbon Tetrachloride	0.489	0.496	-1.4	92	0.00
39 T	Methylcyclohexane	0.595	0.586	1.5	87	0.00
40 TM	Benzene	1.523	1.520	0.2	90	0.00
41 T	Methacrylonitrile	0.340	0.328	3.5	87	0.00
42 TM	1,2-Dichloroethane	0.530	0.541	-2.1	91	0.00
43 T	Isopropyl Acetate	1.040	1.162	-11.7	104	0.00
44 TM	Trichloroethene	0.359	0.364	-1.4	91	0.00
45 C	1,2-Dichloropropane	0.409	0.404	1.2#	89	0.00
46 T	Dibromomethane	0.252	0.255	-1.2	91	0.00
47 T	Bromodichloromethane	0.509	0.529	-3.9	93	0.00
48 T	Methyl methacrylate	0.482	0.510	-5.8	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031422\
 Data File : VN071282.D
 Acq On : 14 Mar 2022 21:39
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 15 04:03:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.009	-12.5	96	0.00
50 S	Toluene-d8	1.270	1.211	4.6	88	0.00
51 T	4-Methyl-2-Pentanone	0.646	0.686	-6.2	93	0.00
52 CM	Toluene	0.938	0.946	-0.9#	88	0.00
53 T	t-1,3-Dichloropropene	0.561	0.573	-2.1	89	0.00
54 T	cis-1,3-Dichloropropene	0.613	0.622	-1.5	89	0.00
55 T	1,1,2-Trichloroethane	0.376	0.379	-0.8	93	0.00
56 T	Ethyl methacrylate	0.614	0.659	-7.3	93	0.00
57 T	1,3-Dichloropropane	0.675	0.675	0.0	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.101	0.207	-105.0#	166#	0.00
59 T	2-Hexanone	0.441	0.501	-13.6	96	0.00
60 T	Dibromochloromethane	0.373	0.396	-6.2	91	0.00
61 T	1,2-Dibromoethane	0.384	0.387	-0.8	90	0.00
62 S	4-Bromofluorobenzene	0.430	0.426	0.9	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	84	0.00
64 T	Tetrachloroethene	0.349	0.335	4.0	87	0.00
65 PM	Chlorobenzene	1.074	1.071	0.3	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.409	-3.8	91	0.00
67 C	Ethyl Benzene	1.899	1.984	-4.5#	90	0.00
68 T	m/p-Xylenes	0.695	0.740	-6.5	89	0.00
69 T	o-Xylene	0.696	0.745	-7.0	90	0.00
70 T	Styrene	1.078	1.201	-11.4	92	0.00
71 P	Bromoform	0.299	0.330	-10.4	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	4.820	4.496	6.7	90	0.00
74 T	N-amyl acetate	1.812	1.904	-5.1	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.655	1.563	5.6	95	0.00
76 T	1,2,3-Trichloropropane	1.407	1.390	1.2	105	0.00
77 T	Bromobenzene	1.124	1.050	6.6	92	0.00
78 T	n-propylbenzene	5.172	5.013	3.1	91	0.00
79 T	2-Chlorotoluene	3.320	3.090	6.9	91	0.00
80 T	1,3,5-Trimethylbenzene	3.750	3.646	2.8	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.463	0.492	-6.3	99	0.00
82 T	4-Chlorotoluene	2.957	2.957	0.0	94	0.00
83 T	tert-Butylbenzene	3.347	3.292	1.6	92	0.00
84 T	1,2,4-Trimethylbenzene	3.592	3.586	0.2	92	0.00
85 T	sec-Butylbenzene	4.448	4.437	0.2	91	0.00
86 T	p-Isopropyltoluene	3.498	3.598	-2.9	92	0.00
87 T	1,3-Dichlorobenzene	1.826	1.781	2.5	94	0.00
88 T	1,4-Dichlorobenzene	1.785	1.699	4.8	94	0.00
89 T	n-Butylbenzene	2.736	2.767	-1.1	93	0.00
90 T	Hexachloroethane	0.696	0.671	3.6	91	0.00
91 T	1,2-Dichlorobenzene	1.756	1.729	1.5	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.280	0.296	-5.7	103	0.00
93 T	1,2,4-Trichlorobenzene	0.797	0.800	-0.4	96	0.00
94 T	Hexachlorobutadiene	0.493	0.461	6.5	88	0.00
95 T	Naphthalene	2.115	2.529	-19.6	115	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031422\
Data File : VN071282.D
Acq On : 14 Mar 2022 21:39
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 15 04:03:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.775	0.819	-5.7	101	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6