

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031423\
 Data File : VN076981.D
 Acq On : 14 Mar 2023 14:44
 Operator : JC\MD
 Sample : PB151382TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB151382TB

Quant Time: Mar 15 04:23:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	272651	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	409648	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	369937	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	167366	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	201367	49.492	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.980%
35) Dibromofluoromethane	8.172	113	131221	48.063	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.120%
50) Toluene-d8	10.572	98	497399	49.433	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.860%
62) 4-Bromofluorobenzene	12.854	95	187247	50.877	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.760%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	25179	17.859	ug/l	91
20) Methylene Chloride	5.290	84	8917	2.460	ug/l #	89
37) Ethyl Acetate	7.572	43	39400	11.668	ug/l	97
43) Isopropyl Acetate	8.695	43	186663	31.314	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031423\
Data File : VN076981.D
Acq On : 14 Mar 2023 14:44
Operator : JC\MD
Sample : PB151382TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB151382TB

Quant Time: Mar 15 04:23:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

