

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031423\
 Data File : VN076994.D
 Acq On : 14 Mar 2023 19:58
 Operator : JC\MD
 Sample : 01905-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MAPLE-2

Quant Time: Mar 15 04:30:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.237	168	245466	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	366334	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	333117	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	149819	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	187500	51.187	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.380%	
35) Dibromofluoromethane	8.172	113	122533	50.188	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.380%	
50) Toluene-d8	10.572	98	430800	47.877	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.760%	
62) 4-Bromofluorobenzene	12.854	95	164169	49.881	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.760%	
Target Compounds						
						Qvalue
16) Acetone	4.449	43	20291	15.986	ug/l #	87
20) Methylene Chloride	5.301	84	8780	2.690	ug/l	92
37) Ethyl Acetate	7.572	43	30273	10.025	ug/l	96
43) Isopropyl Acetate	8.695	43	210631	39.513	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031423\
Data File : VN076994.D
Acq On : 14 Mar 2023 19:58
Operator : JC\MD
Sample : 01905-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MAPLE-2

Quant Time: Mar 15 04:30:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

