

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031424\
 Data File : VN081406.D
 Acq On : 14 Mar 2024 13:41
 Operator : JC\MD
 Sample : P1725-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RT-3860RE

Quant Time: Mar 14 14:26:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	298262	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	548702	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	490027	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	188326	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	231582	53.700	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.400%	
35) Dibromofluoromethane	8.171	113	170201	50.629	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.260%	
50) Toluene-d8	10.570	98	672318	53.499	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.000%	
62) 4-Bromofluorobenzene	12.853	95	209575	47.184	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 94.360%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	51798	33.900	ug/l	89
20) Methylene Chloride	5.288	84	17599	3.920	ug/l #	88
43) Isopropyl Acetate	8.694	43	294195	31.704	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031424\
Data File : VN081406.D
Acq On : 14 Mar 2024 13:41
Operator : JC\MD
Sample : P1725-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RT-3860RE

Quant Time: Mar 14 14:26:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 06 03:12:57 2024
Response via : Initial Calibration

