

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031424\
 Data File : VN081411.D
 Acq On : 14 Mar 2024 15:41
 Operator : JC\MD
 Sample : P1749-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Mar 15 01:14:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	317659	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	576023	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	500051	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	194827	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	236950	51.589	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.180%	
35) Dibromofluoromethane	8.165	113	184514	52.284	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.560%	
50) Toluene-d8	10.571	98	691302	52.401	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.800%	
62) 4-Bromofluorobenzene	12.853	95	218531	46.867	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 93.740%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031424\
Data File : VN081411.D
Acq On : 14 Mar 2024 15:41
Operator : JC\MD
Sample : P1749-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Mar 15 01:14:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 06 03:12:57 2024
Response via : Initial Calibration

