

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031424\
 Data File : VN081424.D
 Acq On : 14 Mar 2024 20:52
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 15 01:19:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	0.663	0.660	0.5	79	0.00
3 P	Chloromethane	0.716	0.632	11.7	81	0.00
4 C	Vinyl Chloride	0.747	0.694	7.1#	85	0.00
5 T	Bromomethane	0.499	0.426	14.6	78	0.00
6 T	Chloroethane	0.513	0.475	7.4	88	0.00
7 T	Trichlorofluoromethane	1.105	1.071	3.1	86	0.00
8 T	Diethyl Ether	0.389	0.382	1.8	86	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.629	0.597	5.1	88	0.01
10 T	Methyl Iodide	0.576	0.618	-7.3	89	0.00
11 T	Tert butyl alcohol	0.126	0.140	-11.1	105	0.01
12 CM	1,1-Dichloroethene	0.583	0.553	5.1#	85	0.01
13 T	Acrolein	0.150	0.122	18.7	72	0.00
14 T	Allyl chloride	0.865	0.832	3.8	88	0.00
15 T	Acrylonitrile	0.330	0.321	2.7	89	0.00
16 T	Acetone	0.256	0.257	-0.4	96	0.00
17 T	Carbon Disulfide	1.663	1.455	12.5	80	0.00
18 T	Methyl Acetate	0.733	0.860	-17.3	101	0.00
19 T	Methyl tert-butyl Ether	1.976	2.034	-2.9	92	0.00
20 T	Methylene Chloride	0.718	0.662	7.8	90	0.00
21 T	trans-1,2-Dichloroethene	0.657	0.622	5.3	88	0.00
22 T	Diisopropyl ether	1.921	2.081	-8.3	92	0.00
23 T	Vinyl Acetate	1.560	1.592	-2.1	87	0.00
24 P	1,1-Dichloroethane	1.176	1.215	-3.3	93	0.00
25 T	2-Butanone	0.429	0.411	4.2	88	0.00
26 T	2,2-Dichloropropane	1.039	0.926	10.9	78	0.00
27 T	cis-1,2-Dichloroethene	0.745	0.726	2.6	91	0.00
28 T	Bromochloromethane	0.503	0.499	0.8	90	0.00
29 T	Tetrahydrofuran	0.293	0.279	4.8	88	0.00
30 C	Chloroform	1.220	1.268	-3.9#	93	0.00
31 T	Cyclohexane	1.095	0.972	11.2	83	0.00
32 T	1,1,1-Trichloroethane	1.078	1.123	-4.2	93	0.00
33 S	1,2-Dichloroethane-d4	0.723	0.670	7.3	76	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	87	0.00
35 S	Dibromofluoromethane	0.306	0.278	9.2	74	0.00
36 T	1,1-Dichloropropene	0.494	0.481	2.6	87	0.00
37 T	Ethyl Acetate	0.483	0.478	1.0	87	0.00
38 T	Carbon Tetrachloride	0.497	0.514	-3.4	89	0.00
39 T	Methylcyclohexane	0.569	0.524	7.9	79	0.00
40 TM	Benzene	1.475	1.476	-0.1	90	0.00
41 T	Methacrylonitrile	0.273	0.276	-1.1	94	0.00
42 TM	1,2-Dichloroethane	0.520	0.548	-5.4	94	0.00
43 T	Isopropyl Acetate	0.846	0.812	4.0	87	0.00
44 TM	Trichloroethene	0.372	0.373	-0.3	89	0.00
45 C	1,2-Dichloropropane	0.371	0.389	-4.9#	94	0.00
46 T	Dibromomethane	0.261	0.271	-3.8	91	0.00
47 T	Bromodichloromethane	0.507	0.547	-7.9	95	0.00
48 T	Methyl methacrylate	0.374	0.386	-3.2	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031424\
 Data File : VN081424.D
 Acq On : 14 Mar 2024 20:52
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 15 01:19:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	86	0.00
50 S	Toluene-d8	1.145	1.052	8.1	74	0.00
51 T	4-Methyl-2-Pentanone	0.491	0.501	-2.0	90	0.00
52 CM	Toluene	0.891	0.925	-3.8#	90	0.00
53 T	t-1,3-Dichloropropene	0.549	0.579	-5.5	90	0.00
54 T	cis-1,3-Dichloropropene	0.592	0.618	-4.4	90	0.00
55 T	1,1,2-Trichloroethane	0.353	0.368	-4.2	93	0.00
56 T	Ethyl methacrylate	0.507	0.554	-9.3	90	0.00
57 T	1,3-Dichloropropane	0.617	0.637	-3.2	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.273	0.291	-6.6	87	0.00
59 T	2-Hexanone	0.367	0.372	-1.4	89	0.00
60 T	Dibromochloromethane	0.364	0.395	-8.5	93	0.00
61 T	1,2-Dibromoethane	0.367	0.362	1.4	87	0.00
62 S	4-Bromofluorobenzene	0.405	0.388	4.2	77	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.391	0.392	-0.3	92	0.00
65 PM	Chlorobenzene	1.054	1.066	-1.1	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.386	0.396	-2.6	92	0.00
67 C	Ethyl Benzene	1.898	1.963	-3.4#	90	0.00
68 T	m/p-Xylenes	0.714	0.735	-2.9	90	0.00
69 T	o-Xylene	0.692	0.716	-3.5	90	0.00
70 T	Styrene	1.097	1.214	-10.7	91	0.00
71 P	Bromoform	0.268	0.278	-3.7	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	4.037	4.118	-2.0	89	0.00
74 T	N-amyl acetate	1.694	1.731	-2.2	88	0.00
75 P	1,1,2,2-Tetrachloroethane	1.246	1.152	7.5	88	0.00
76 T	1,2,3-Trichloropropane	1.181	1.133	4.1	91	0.00
77 T	Bromobenzene	0.957	0.914	4.5	90	0.00
78 T	n-propylbenzene	4.775	4.972	-4.1	88	0.00
79 T	2-Chlorotoluene	2.919	2.957	-1.3	91	0.00
80 T	1,3,5-Trimethylbenzene	3.338	3.484	-4.4	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.423	0.408	3.5	76	0.00
82 T	4-Chlorotoluene	2.883	2.927	-1.5	91	0.00
83 T	tert-Butylbenzene	2.861	2.894	-1.2	89	0.00
84 T	1,2,4-Trimethylbenzene	3.381	3.526	-4.3	90	0.00
85 T	sec-Butylbenzene	4.139	4.216	-1.9	87	0.00
86 T	p-Isopropyltoluene	3.320	3.518	-6.0	89	0.00
87 T	1,3-Dichlorobenzene	1.808	1.720	4.9	90	0.00
88 T	1,4-Dichlorobenzene	1.807	1.708	5.5	89	0.00
89 T	n-Butylbenzene	3.044	3.114	-2.3	87	0.00
90 T	Hexachloroethane	0.581	0.622	-7.1	92	0.00
91 T	1,2-Dichlorobenzene	1.734	1.656	4.5	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.267	0.241	9.7	82	0.00
93 T	1,2,4-Trichlorobenzene	0.927	0.901	2.8	85	0.00
94 T	Hexachlorobutadiene	0.383	0.335	12.5	81	0.00
95 T	Naphthalene	3.299	3.267	1.0	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031424\
Data File : VN081424.D
Acq On : 14 Mar 2024 20:52
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 15 01:19:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 06 03:12:57 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.940	0.924	1.7	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6