

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031424\
 Data File : VN081412.D
 Acq On : 14 Mar 2024 16:05
 Operator : JC\MD
 Sample : P1747-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-01

Quant Time: Mar 15 01:14:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.230	168	303621	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	551652	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	483758	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	192317	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	230511	52.508	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.020%	
35) Dibromofluoromethane	8.171	113	177254	52.445	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.900%	
50) Toluene-d8	10.571	98	659893	52.230	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.460%	
62) 4-Bromofluorobenzene	12.853	95	209487	46.912	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 93.820%	
Target Compounds						
						Qvalue
30) Chloroform	7.971	83	229758	31.011	ug/l	100
47) Bromodichloromethane	9.894	83	16265	2.907	ug/l	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031424\
Data File : VN081412.D
Acq On : 14 Mar 2024 16:05
Operator : JC\MD
Sample : P1747-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-01

Quant Time: Mar 15 01:14:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 06 03:12:57 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN081412.D
Acq: 14 Mar 2024 16:05




Instrument :
MSVOA_N
ClientSampleId :
MW-01


Tgt Ion:168 Resp: 303621
Ion Ratio Lower Upper
168 100
99 63.3 59.9 89.9



#30
Chloroform
Concen: 31.011 ug/l
RT: 7.971 min Scan# 1023
Delta R.T. 0.000 min
Lab File: VN081412.D
Acq: 14 Mar 2024 16:05

Tgt Ion: 83 Resp: 229758
Ion Ratio Lower Upper
83 100
85 63.8 51.0 76.6





#33

1,2-Dichloroethane-d4

Concen: 52.508 ug/l

RT: 8.583 min Scan# 1126

Delta R.T. 0.006 min

Lab File: VN081412.D

Acq: 14 Mar 2024 16:05

Instrument :

MSVOA_N

ClientSampleId :





MW-01

Tgt Ion: 65 Resp: 230511

Ion Ratio Lower Upper

65 100

67 51.8 0.0 102.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. 0.000 min

Lab File: VN081412.D

Acq: 14 Mar 2024 16:05




Tgt Ion: 114 Resp: 551652


Ion Ratio Lower Upper

114 100

63 22.2 0.0 48.0

88 16.5 0.0 34.8





#35

Dibromofluoromethane

Concen: 52.445 ug/l

RT: 8.171 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN081412.D

Acq: 14 Mar 2024 16:05

Instrument :

MSVOA_N

ClientSampleId :

MW-01





Tgt Ion:113 Resp: 177254

Ion Ratio Lower Upper

113 100

111 101.9 82.2 123.4

192 15.3 12.5 18.7



#47

Bromodichloromethane

Concen: 2.907 ug/l

RT: 9.894 min Scan# 1350

Delta R.T. 0.006 min

Lab File: VN081412.D

Acq: 14 Mar 2024 16:05




Tgt Ion: 83 Resp: 16265


Ion Ratio Lower Upper

83 100

85 58.6 52.3 78.5

127 5.9 4.6 7.0









#50
Toluene-d8
Concen: 52.230 ug/l
RT: 10.571 min Scan# 1465
Delta R.T. 0.000 min
Lab File: VN081412.D
Acq: 14 Mar 2024 16:05




Instrument :
MSVOA_N
ClientSampleId :
MW-01


Tgt Ion: 98 Resp: 659893
Ion Ratio Lower Upper
98 100
100 64.9 51.4 77.0



#62
4-Bromofluorobenzene
Concen: 46.912 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.000 min
Lab File: VN081412.D
Acq: 14 Mar 2024 16:05

Tgt Ion: 95 Resp: 209487
Ion Ratio Lower Upper
95 100
174 69.0 0.0 120.4
176 65.9 0.0 121.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.871 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN081412.D

Acq: 14 Mar 2024 16:05

Instrument :

MSVOA_N

ClientSampleId :

MW-01





Tgt Ion:117 Resp: 483758

Ion Ratio Lower Upper

117 100

82 57.9 52.7 79.1

119 33.6 25.3 37.9



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. 0.000 min

Lab File: VN081412.D

Acq: 14 Mar 2024 16:05

Tgt Ion:152 Resp: 192317

Ion Ratio Lower Upper

152 100

115 63.0 32.3 96.8

150 152.6 0.0 339.8

