

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031425\
 Data File : VN085993.D
 Acq On : 14 Mar 2025 16:45
 Operator : JC\MD
 Sample : VSTDC0050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDC0050EC

Quant Time: Mar 15 00:05:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	171602	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	273448	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	248078	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	133054	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	83945	37.732	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	75.460%
35) Dibromofluoromethane	8.165	113	68137	38.080	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	76.160%
50) Toluene-d8	10.565	98	244917	37.621	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	75.240%#
62) 4-Bromofluorobenzene	12.847	95	92696	43.217	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	86.440%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	105628	45.254	ug/l	95
3) Chloromethane	2.359	50	84263	35.976	ug/l	96
4) Vinyl Chloride	2.512	62	89373	36.039	ug/l	98
5) Bromomethane	2.948	94	54028	34.072	ug/l	87
6) Chloroethane	3.112	64	53047	32.294	ug/l	89
7) Trichlorofluoromethane	3.495	101	157099	43.754	ug/l	99
8) Diethyl Ether	3.959	74	50696	43.638	ug/l	96
9) 1,1,2-Trichlorotrifluo...	4.371	101	87016	41.926	ug/l	100
10) Methyl Iodide	4.589	142	115648	45.204	ug/l	95
11) Tert butyl alcohol	5.512	59	59536	217.579	ug/l	98
12) 1,1-Dichloroethene	4.342	96	81556	43.338	ug/l	97
13) Acrolein	4.177	56	75690	186.532	ug/l	100
14) Allyl chloride	5.030	41	103201	41.770	ug/l	96
15) Acrylonitrile	5.712	53	191076	214.801	ug/l	99
16) Acetone	4.424	43	140261	199.247	ug/l	98
17) Carbon Disulfide	4.712	76	233105	41.796	ug/l	100
18) Methyl Acetate	5.018	43	82277	33.981	ug/l	96
19) Methyl tert-butyl Ether	5.794	73	276679	49.477	ug/l	97
20) Methylene Chloride	5.277	84	93290	41.209	ug/l	98
21) trans-1,2-Dichloroethene	5.783	96	86160	42.521	ug/l	94
22) Diisopropyl ether	6.671	45	235204	42.130	ug/l	96
23) Vinyl Acetate	6.600	43	836575	218.850	ug/l	97
24) 1,1-Dichloroethane	6.565	63	159383	41.996	ug/l	98
25) 2-Butanone	7.483	43	213105	205.770	ug/l	98
26) 2,2-Dichloropropane	7.489	77	152426	45.167	ug/l	98
27) cis-1,2-Dichloroethene	7.489	96	101207	43.076	ug/l	99
28) Bromochloromethane	7.812	49	60653	38.884	ug/l	95
29) Tetrahydrofuran	7.836	42	144285	215.749	ug/l	92
30) Chloroform	7.965	83	173735	43.805	ug/l	100
31) Cyclohexane	8.259	56	128723	40.509	ug/l	98
32) 1,1,1-Trichloroethane	8.165	97	162885	46.174	ug/l	97
36) 1,1-Dichloropropene	8.371	75	117235	46.295	ug/l	99
37) Ethyl Acetate	7.553	43	87948	42.202	ug/l #	97
38) Carbon Tetrachloride	8.359	117	151549	50.366	ug/l	93
39) Methylcyclohexane	9.600	83	117598	47.270	ug/l	96
40) Benzene	8.606	78	371456	45.766	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031425\
 Data File : VN085993.D
 Acq On : 14 Mar 2025 16:45
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Mar 15 00:05:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	48022	43.106	ug/l	89
42) 1,2-Dichloroethane	8.671	62	125546	47.488	ug/l	99
43) Isopropyl Acetate	8.683	43	161236	44.020	ug/l	96
44) Trichloroethene	9.347	130	87704	44.010	ug/l	97
45) 1,2-Dichloropropane	9.618	63	87741	44.785	ug/l	93
46) Dibromomethane	9.706	93	65589	47.463	ug/l	99
47) Bromodichloromethane	9.888	83	139486	47.179	ug/l	100
48) Methyl methacrylate	9.677	41	72736	46.996	ug/l	94
49) 1,4-Dioxane	9.688	88	31233	1061.219	ug/l #	94
51) 4-Methyl-2-Pentanone	10.441	43	467884	232.866	ug/l	95
52) Toluene	10.629	92	241448	50.749	ug/l	98
53) t-1,3-Dichloropropene	10.835	75	139392	50.987	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	145823	48.642	ug/l	98
55) 1,1,2-Trichloroethane	11.012	97	87615	46.317	ug/l	97
56) Ethyl methacrylate	10.871	69	133550	49.200	ug/l	95
57) 1,3-Dichloropropane	11.159	76	150019	47.508	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	260575	247.060	ug/l	96
59) 2-Hexanone	11.194	43	337664	244.286	ug/l	95
60) Dibromochloromethane	11.359	129	111245	50.504	ug/l	98
61) 1,2-Dibromoethane	11.465	107	88474	49.502	ug/l	98
64) Tetrachloroethene	11.100	164	83686	43.544	ug/l	97
65) Chlorobenzene	11.888	112	256915	45.207	ug/l	94
66) 1,1,1,2-Tetrachloroethane	11.959	131	95886	46.941	ug/l	97
67) Ethyl Benzene	11.959	91	447268	50.242	ug/l	99
68) m/p-Xylenes	12.071	106	354743	104.906	ug/l	97
69) o-Xylene	12.394	106	169179	52.649	ug/l	98
70) Styrene	12.406	104	289792	48.607	ug/l	98
71) Bromoform	12.576	173	78490	51.711	ug/l #	98
73) Isopropylbenzene	12.694	105	424070	46.522	ug/l	100
74) N-amyl acetate	12.488	43	138967	43.231	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.935	83	124086	39.559	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	164081	55.516	ug/l	57
77) Bromobenzene	12.976	156	107239	43.545	ug/l	98
78) n-propylbenzene	13.035	91	495957	47.508	ug/l	100
79) 2-Chlorotoluene	13.123	91	316622	45.394	ug/l	100
80) 1,3,5-Trimethylbenzene	13.170	105	369726	49.792	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	43187	42.634	ug/l	96
82) 4-Chlorotoluene	13.218	91	322574	46.650	ug/l	99
83) tert-Butylbenzene	13.435	119	302967	48.520	ug/l	96
84) 1,2,4-Trimethylbenzene	13.476	105	377701	51.413	ug/l	99
85) sec-Butylbenzene	13.612	105	413853	46.910	ug/l	98
86) p-Isopropyltoluene	13.729	119	358551	44.790	ug/l	98
87) 1,3-Dichlorobenzene	13.729	146	201826	43.742	ug/l	99
88) 1,4-Dichlorobenzene	13.806	146	200569	42.412	ug/l	99
89) n-Butylbenzene	14.053	91	279807	44.671	ug/l	99
90) Hexachloroethane	14.329	117	70023	41.909	ug/l	94
91) 1,2-Dichlorobenzene	14.106	146	195290	44.015	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	24787	43.033	ug/l	99
93) 1,2,4-Trichlorobenzene	15.388	180	94648	44.768	ug/l	97
94) Hexachlorobutadiene	15.500	225	50783	39.523	ug/l	99
95) Naphthalene	15.635	128	272761	46.992	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	93144	43.884	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031425\
Data File : VN085993.D
Acq On : 14 Mar 2025 16:45
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Quant Time: Mar 15 00:05:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031425\
 Data File : VN085993.D
 Acq On : 14 Mar 2025 16:45
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Mar 15 00:05:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:168 Resp: 171602
Ion Ratio Lower Upper
168 100
99 58.9 47.9 71.9



#2
Dichlorodifluoromethane
Concen: 45.254 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion: 85 Resp: 105628
Ion Ratio Lower Upper
85 100
87 29.8 16.4 49.2









#3
Chloromethane
Concen: 35.976 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 50 Resp: 84263
Ion Ratio Lower Upper
50 100
52 31.0 26.7 40.1



#4
Vinyl Chloride
Concen: 36.039 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion: 62 Resp: 89373
Ion Ratio Lower Upper
62 100
64 31.2 25.9 38.9









#5
Bromomethane
Concen: 34.072 ug/l
RT: 2.948 min Scan# 1
Delta R.T. -0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 94 Resp: 54028
Ion Ratio Lower Upper
94 100
96 101.5 71.7 107.5



#6
Chloroethane
Concen: 32.294 ug/l
RT: 3.112 min Scan# 197
Delta R.T. -0.012 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion: 64 Resp: 53047
Ion Ratio Lower Upper
64 100
66 27.5 27.0 40.6









#7
 Trichlorofluoromethane
 Concen: 43.754 ug/l
 RT: 3.495 min Scan# 262
 Delta R.T. -0.006 min
 Lab File: VN085993.D
 Acq: 14 Mar 2025 16:45




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:101 Resp: 157099
 Ion Ratio Lower Upper
 101 100
 103 64.6 52.2 78.4



#8
 Diethyl Ether
 Concen: 43.638 ug/l
 RT: 3.959 min Scan# 341
 Delta R.T. 0.000 min
 Lab File: VN085993.D
 Acq: 14 Mar 2025 16:45

Tgt Ion: 74 Resp: 50696
 Ion Ratio Lower Upper
 74 100
 45 81.2 42.5 127.6









#9
1,1,2-Trichlorotrifluoroethane
Concen: 41.926 ug/l
RT: 4.371 min Scan# 41
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 87016
Ion Ratio Lower Upper
101 100
85 44.8 35.7 53.5
151 76.3 61.4 92.2



#10
Methyl Iodide
Concen: 45.204 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion:142 Resp: 115648
Ion Ratio Lower Upper
142 100
127 51.7 37.6 56.4
141 14.3 11.4 17.2









#11
Tert butyl alcohol
Concen: 217.579 ug/l
RT: 5.512 min Scan# 604
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 59 Resp: 59536
Ion Ratio Lower Upper
59 100
57 11.0 8.3 12.5



#12
1,1-Dichloroethene
Concen: 43.338 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion: 96 Resp: 81556
Ion Ratio Lower Upper
96 100
61 157.6 123.3 184.9
98 64.6 48.6 72.8









#13
Acrolein
Concen: 186.532 ug/l
RT: 4.177 min Scan# 31
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 75690
Ion Ratio Lower Upper
56 100
55 69.4 55.3 82.9



#14
Allyl chloride
Concen: 41.770 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.012 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion: 41 Resp: 103201
Ion Ratio Lower Upper
41 100
39 66.7 57.4 86.2
76 45.6 36.2 54.2









#15
Acrylonitrile
Concen: 214.801 ug/l
RT: 5.712 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 53 Resp: 191076
Ion Ratio Lower Upper
53 100
52 81.1 65.8 98.6
51 35.4 28.6 42.8



#16
Acetone
Concen: 199.247 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion: 43 Resp: 140261
Ion Ratio Lower Upper
43 100
58 35.8 27.5 41.3









#17
Carbon Disulfide
Concen: 41.796 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 76 Resp: 233105
Ion Ratio Lower Upper
76 100
78 9.3 7.4 11.0



#18
Methyl Acetate
Concen: 33.981 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion: 43 Resp: 82277
Ion Ratio Lower Upper
43 100
74 29.8 22.2 33.4









#19
Methyl tert-butyl Ether
Concen: 49.477 ug/l
RT: 5.794 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 73 Resp: 276679
Ion Ratio Lower Upper
73 100
57 20.5 17.6 26.4



#20
Methylene Chloride
Concen: 41.209 ug/l
RT: 5.277 min Scan# 565
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion: 84 Resp: 93290
Ion Ratio Lower Upper
84 100
49 108.4 88.4 132.6
51 31.3 26.6 40.0
86 62.2 51.0 76.6









#21
trans-1,2-Dichloroethene
Concen: 42.521 ug/l
RT: 5.783 min Scan# 651
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 96 Resp: 86160
Ion Ratio Lower Upper
96 100
61 133.7 112.6 168.8
98 61.1 53.3 79.9



#22
Diisopropyl ether
Concen: 42.130 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion: 45 Resp: 235204
Ion Ratio Lower Upper
45 100
43 55.8 42.6 64.0
87 34.4 25.8 38.8
59 14.3 9.9 14.9









#23
 Vinyl Acetate
 Concen: 218.850 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. 0.006 min
 Lab File: VN085993.D
 Acq: 14 Mar 2025 16:45




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 43 Resp: 836575
 Ion Ratio Lower Upper
 43 100
 86 13.3 9.8 14.8



#24
 1,1-Dichloroethane
 Concen: 41.996 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. 0.000 min
 Lab File: VN085993.D
 Acq: 14 Mar 2025 16:45

Tgt Ion: 63 Resp: 159383
 Ion Ratio Lower Upper
 63 100
 98 7.6 3.5 10.4
 100 5.2 2.4 7.2









#25
2-Butanone
Concen: 205.770 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.012 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 213105
Ion Ratio Lower Upper
43 100
72 30.8 23.6 35.4



#26
2,2-Dichloropropane
Concen: 45.167 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion: 77 Resp: 152426
Ion Ratio Lower Upper
77 100
97 21.7 11.3 33.8





#27

cis-1,2-Dichloroethene

Concen: 43.076 ug/l

RT: 7.489 min Scan# 941

Delta R.T. 0.006 min

Lab File: VN085993.D

Acq: 14 Mar 2025 16:45

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 96 Resp: 101207

Ion Ratio Lower Upper

96 100

61 142.0 0.0 280.6

98 65.1 0.0 128.6



#28

Bromochloromethane

Concen: 38.884 ug/l

RT: 7.812 min Scan# 996

Delta R.T. 0.006 min

Lab File: VN085993.D

Acq: 14 Mar 2025 16:45




Tgt Ion: 49 Resp: 60653


Ion Ratio Lower Upper

49 100

129 1.8 0.0 3.8

130 89.6 68.0 102.0





#29

Tetrahydrofuran

Concen: 215.749 ug/l

RT: 7.836 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN085993.D

Acq: 14 Mar 2025 16:45

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 42 Resp: 144285

Ion Ratio Lower Upper

42 100

72 56.1 40.0 60.0

71 50.4 37.0 55.4



#30

Chloroform

Concen: 43.805 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.006 min

Lab File: VN085993.D




Acq: 14 Mar 2025 16:45


Tgt Ion: 83 Resp: 173735

Ion Ratio Lower Upper

83 100

85 63.7 50.9 76.3









#31
Cyclohexane
Concen: 40.509 ug/l
RT: 8.259 min Scan# 1071
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 128723
Ion Ratio Lower Upper
56 100
69 31.4 25.8 38.6
84 94.7 74.5 111.7



#32
1,1,1-Trichloroethane
Concen: 46.174 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion: 97 Resp: 162885
Ion Ratio Lower Upper
97 100
99 65.5 50.4 75.6
61 43.0 35.5 53.3









#33
 1,2-Dichloroethane-d4
 Concen: 37.732 ug/l
 RT: 8.577 min Scan# 1125
 Delta R.T. 0.006 min
 Lab File: VN085993.D
 Acq: 14 Mar 2025 16:45




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 65 Resp: 83945
 Ion Ratio Lower Upper
 65 100
 67 51.1 0.0 106.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 9.100 min Scan# 1215
 Delta R.T. 0.006 min
 Lab File: VN085993.D
 Acq: 14 Mar 2025 16:45

Tgt Ion: 114 Resp: 273448
 Ion Ratio Lower Upper
 114 100
 63 18.7 0.0 39.2
 88 15.8 0.0 30.0





#35

Dibromofluoromethane

Concen: 38.080 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN085993.D

Acq: 14 Mar 2025 16:45

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:113 Resp: 68137

Ion Ratio Lower Upper

113 100

111 103.9 82.2 123.4

192 19.3 14.7 22.1



#36

1,1-Dichloropropene

Concen: 46.295 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.006 min

Lab File: VN085993.D

Acq: 14 Mar 2025 16:45




Tgt Ion: 75 Resp: 117235


Ion Ratio Lower Upper

75 100

110 35.4 17.3 52.0

77 30.5 24.6 37.0









#37
Ethyl Acetate
Concen: 42.202 ug/l
RT: 7.553 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 87948
Ion Ratio Lower Upper
43 100
61 15.2 12.2 18.2
70 14.4 9.5 14.3#



#38
Carbon Tetrachloride
Concen: 50.366 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion: 117 Resp: 151549
Ion Ratio Lower Upper
117 100
119 100.1 73.8 110.6
121 29.2 24.3 36.5









#39
Methylcyclohexane
Concen: 47.270 ug/l
RT: 9.600 min Scan# 111
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 117598
Ion Ratio Lower Upper
83 100
55 77.1 57.5 86.3
98 50.1 39.2 58.8



#40
Benzene
Concen: 45.766 ug/l
RT: 8.606 min Scan# 1131
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion: 78 Resp: 371456
Ion Ratio Lower Upper
78 100
77 23.1 18.3 27.5









#41
Methacrylonitrile
Concen: 43.106 ug/l
RT: 7.777 min Scan# 91
Delta R.T. 0.012 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	41	Resp:	48022
Ion	Ratio	Lower	Upper
41	100		
39	56.7	42.0	63.0
67	94.7	64.2	96.4
52	34.9	24.7	37.1



#42
1,2-Dichloroethane
Concen: 47.488 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion:	62	Resp:	125546
Ion	Ratio	Lower	Upper
62	100		
98	9.5	0.0	19.6





#43

Isopropyl Acetate

Concen: 44.020 ug/l

RT: 8.683 min Scan# 1144

Delta R.T. -0.000 min

Lab File: VN085993.D

Acq: 14 Mar 2025 16:45

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 43 Resp: 161236

Ion Ratio Lower Upper

43 100

61 30.4 22.6 34.0

87 16.3 11.8 17.8



#44

Trichloroethene

Concen: 44.010 ug/l

RT: 9.347 min Scan# 1257

Delta R.T. 0.000 min

Lab File: VN085993.D




Acq: 14 Mar 2025 16:45


Tgt Ion:130 Resp: 87704

Ion Ratio Lower Upper

130 100

95 98.6 0.0 202.6





#45

1,2-Dichloropropane

Concen: 44.785 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. 0.000 min

Lab File: VN085993.D

Acq: 14 Mar 2025 16:45

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 63 Resp: 87741

Ion Ratio Lower Upper

63 100

65 35.1 25.1 37.7



#46

Dibromomethane

Concen: 47.463 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.006 min

Lab File: VN085993.D

Acq: 14 Mar 2025 16:45




Tgt Ion: 93 Resp: 65589


Ion Ratio Lower Upper

93 100

95 80.8 66.7 100.1

174 96.3 77.2 115.8









#47
Bromodichloromethane
Concen: 47.179 ug/l
RT: 9.888 min Scan# 1348
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 139486
Ion Ratio Lower Upper
83 100
85 63.3 50.5 75.7
127 9.7 7.2 10.8



#48
Methyl methacrylate
Concen: 46.996 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion: 41 Resp: 72736
Ion Ratio Lower Upper
41 100
69 101.7 76.0 114.0
39 56.7 42.3 63.5





#49

1,4-Dioxane

Concen: 1061.219 ug/l

RT: 9.688 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN085993.D

Acq: 14 Mar 2025 16:45

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 88 Resp: 31233

Ion Ratio Lower Upper

88 100

43 29.1 18.6 27.8#

58 67.5 56.7 85.1



#50

Toluene-d8

Concen: 37.621 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.006 min

Lab File: VN085993.D




Acq: 14 Mar 2025 16:45


Tgt Ion: 98 Resp: 244917

Ion Ratio Lower Upper

98 100

100 65.9 52.1 78.1









#51
4-Methyl-2-Pentanone
Concen: 232.866 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 467884
Ion Ratio Lower Upper
43 100
58 46.4 34.6 52.0



#52
Toluene
Concen: 50.749 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion: 92 Resp: 241448
Ion Ratio Lower Upper
92 100
91 167.7 136.9 205.3









#53
 t-1,3-Dichloropropene
 Concen: 50.987 ug/l
 RT: 10.835 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: VN085993.D
 Acq: 14 Mar 2025 16:45




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 75 Resp: 139392
 Ion Ratio Lower Upper
 75 100
 77 31.0 25.4 38.0



#54
 cis-1,3-Dichloropropene
 Concen: 48.642 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. 0.006 min
 Lab File: VN085993.D
 Acq: 14 Mar 2025 16:45

Tgt Ion: 75 Resp: 145823
 Ion Ratio Lower Upper
 75 100
 77 32.9 26.1 39.1
 39 40.0 33.4 50.0





#55

1,1,2-Trichloroethane

Concen: 46.317 ug/l

RT: 11.012 min Scan# 1540

Delta R.T. -0.000 min

Lab File: VN085993.D

Acq: 14 Mar 2025 16:45

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 97 Resp: 87615





Ion Ratio Lower Upper

97 100

83 87.0 66.1 99.1

85 56.1 44.7 67.1

99 60.1 49.7 74.5



#56

Ethyl methacrylate

Concen: 49.200 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN085993.D

Acq: 14 Mar 2025 16:45




Tgt Ion: 69 Resp: 133550


Ion Ratio Lower Upper

69 100

41 52.5 46.0 69.0

39 28.9 22.0 33.0





#57

1,3-Dichloropropane

Concen: 47.508 ug/l

RT: 11.159 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN085993.D

Acq: 14 Mar 2025 16:45

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 76 Resp: 150019

Ion Ratio Lower Upper

76 100

78 32.4 25.9 38.9



#58

2-Chloroethyl Vinyl ether

Concen: 247.060 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.006 min

Lab File: VN085993.D




Acq: 14 Mar 2025 16:45


Tgt Ion: 63 Resp: 260575

Ion Ratio Lower Upper

63 100

106 30.8 22.7 34.1









#59
2-Hexanone
Concen: 244.286 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 337664
Ion Ratio Lower Upper
43 100
58 63.4 29.8 89.5



#60
Dibromochloromethane
Concen: 50.504 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion:129 Resp: 111245
Ion Ratio Lower Upper
129 100
127 76.6 39.1 117.5









#61
1,2-Dibromoethane
Concen: 49.502 ug/l
RT: 11.465 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:107 Resp: 88474
Ion Ratio Lower Upper
107 100
109 93.4 76.0 114.0



#62
4-Bromofluorobenzene
Concen: 43.217 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion: 95 Resp: 92696
Ion Ratio Lower Upper
95 100
174 80.1 0.0 152.4
176 75.6 0.0 146.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:117 Resp: 248078
Ion Ratio Lower Upper
117 100
82 54.0 45.7 68.5
119 32.2 26.2 39.2



#64
Tetrachloroethene
Concen: 43.544 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion:164 Resp: 83686
Ion Ratio Lower Upper
164 100
166 126.1 107.2 160.8
129 99.9 80.7 121.1
131 92.0 74.6 112.0









#65
Chlorobenzene
Concen: 45.207 ug/l
RT: 11.888 min Scan# 112
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:112 Resp: 256915
Ion Ratio Lower Upper
112 100
114 33.7 24.3 36.5



#66
1,1,1,2-Tetrachloroethane
Concen: 46.941 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion:131 Resp: 95886
Ion Ratio Lower Upper
131 100
133 96.6 47.7 143.1
119 62.2 33.3 99.8









#67
Ethyl Benzene
Concen: 50.242 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 447268
Ion Ratio Lower Upper
91 100
106 30.4 24.7 37.1



#68
m/p-Xylenes
Concen: 104.906 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion:106 Resp: 354743
Ion Ratio Lower Upper
106 100
91 209.1 163.3 244.9









#69
o-Xylene
Concen: 52.649 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:106 Resp: 169179
Ion Ratio Lower Upper
106 100
91 217.0 107.1 321.4



#70
Styrene
Concen: 48.607 ug/l
RT: 12.406 min Scan# 1777
Delta R.T. -0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion:104 Resp: 289792
Ion Ratio Lower Upper
104 100
78 51.5 39.0 58.6
103 54.1 43.3 64.9









#71
Bromoform
Concen: 51.711 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:173 Resp: 78490
Ion Ratio Lower Upper
173 100
175 48.6 25.1 75.4
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion:152 Resp: 133054
Ion Ratio Lower Upper
152 100
115 63.1 30.4 91.3
150 171.2 0.0 345.4





#73

Isopropylbenzene

Concen: 46.522 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. 0.006 min

Lab File: VN085993.D

Acq: 14 Mar 2025 16:45

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion:105 Resp: 424070

Ion Ratio Lower Upper

105 100

120 26.2 13.0 39.0



#74

N-ethyl acetate

Concen: 43.231 ug/l

RT: 12.488 min Scan# 1791

Delta R.T. 0.000 min

Lab File: VN085993.D

Acq: 14 Mar 2025 16:45

Tgt Ion: 43 Resp: 138967




Ion Ratio Lower Upper


43 100

70 53.4 41.0 61.4

55 31.0 24.6 36.8

61 29.2 21.6 32.4









#75
 1,1,2,2-Tetrachloroethane
 Concen: 39.559 ug/l
 RT: 12.935 min Scan# 1867
 Delta R.T. 0.000 min
 Lab File: VN085993.D
 Acq: 14 Mar 2025 16:45




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 83 Resp: 124086
 Ion Ratio Lower Upper
 83 100
 131 10.6 5.0 15.0
 85 64.7 32.2 96.6



#76
 1,2,3-Trichloropropane
 Concen: 55.516 ug/l
 RT: 12.994 min Scan# 1877
 Delta R.T. 0.006 min
 Lab File: VN085993.D
 Acq: 14 Mar 2025 16:45

Tgt Ion: 75 Resp: 164081
 Ion Ratio Lower Upper
 75 100
 77 134.9 100.0 300.0









#77
Bromobenzene
Concen: 43.545 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:156 Resp: 107239
Ion Ratio Lower Upper
156 100
77 206.4 104.1 312.3
158 98.7 48.1 144.4



#78
n-propylbenzene
Concen: 47.508 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion: 91 Resp: 495957
Ion Ratio Lower Upper
91 100
120 22.7 11.3 33.8









#79
2-Chlorotoluene
Concen: 45.394 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 316622
Ion Ratio Lower Upper
91 100
126 32.5 16.4 49.1



#80
1,3,5-Trimethylbenzene
Concen: 49.792 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion: 105 Resp: 369726
Ion Ratio Lower Upper
105 100
120 48.1 24.3 72.9









#81
trans-1,4-Dichloro-2-butene
Concen: 42.634 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.006 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 75 Resp: 43187
Ion Ratio Lower Upper
75 100
53 90.8 74.2 111.4
89 49.3 34.9 52.3



#82
4-Chlorotoluene
Concen: 46.650 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion: 91 Resp: 322574
Ion Ratio Lower Upper
91 100
126 32.1 16.4 49.4









#83
 tert-Butylbenzene
 Concen: 48.520 ug/l
 RT: 13.435 min Scan# 1952
 Delta R.T. 0.000 min
 Lab File: VN085993.D
 Acq: 14 Mar 2025 16:45




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:119 Resp: 302967
 Ion Ratio Lower Upper
 119 100
 91 69.8 32.5 97.5
 134 27.0 13.3 39.9



#84
 1,2,4-Trimethylbenzene
 Concen: 51.413 ug/l
 RT: 13.476 min Scan# 1959
 Delta R.T. 0.000 min
 Lab File: VN085993.D
 Acq: 14 Mar 2025 16:45

Tgt Ion:105 Resp: 377701
 Ion Ratio Lower Upper
 105 100
 120 44.0 22.4 67.3





#85

sec-Butylbenzene

Concen: 46.910 ug/l

RT: 13.612 min Scan# 1982

Delta R.T. -0.000 min

Lab File: VN085993.D

Acq: 14 Mar 2025 16:45

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion:105 Resp: 413853

Ion Ratio Lower Upper

105 100

134 20.4 9.8 29.4



#86

p-Isopropyltoluene

Concen: 44.790 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.006 min

Lab File: VN085993.D

Acq: 14 Mar 2025 16:45




Tgt Ion:119 Resp: 358551


Ion Ratio Lower Upper

119 100

134 25.9 13.0 39.0

91 24.9 11.7 35.1









#87
1,3-Dichlorobenzene
Concen: 43.742 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:146 Resp: 201826
Ion Ratio Lower Upper
146 100
111 42.5 21.1 63.3
148 62.8 32.1 96.5



#88
1,4-Dichlorobenzene
Concen: 42.412 ug/l
RT: 13.806 min Scan# 2015
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion:146 Resp: 200569
Ion Ratio Lower Upper
146 100
111 42.4 20.4 61.3
148 64.1 31.8 95.4





#89

n-Butylbenzene

Concen: 44.671 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN085993.D

Acq: 14 Mar 2025 16:45

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 91 Resp: 279807

Ion Ratio Lower Upper

91 100

92 53.3 27.1 81.3

134 25.9 12.7 38.1



#90

Hexachloroethane

Concen: 41.909 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. 0.000 min

Lab File: VN085993.D




Acq: 14 Mar 2025 16:45


Tgt Ion: 117 Resp: 70023

Ion Ratio Lower Upper

117 100

201 76.1 35.6 106.9









#91
 1,2-Dichlorobenzene
 Concen: 44.015 ug/l
 RT: 14.106 min Scan# 2066
 Delta R.T. 0.006 min
 Lab File: VN085993.D
 Acq: 14 Mar 2025 16:45




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:146 Resp: 195290
 Ion Ratio Lower Upper
 146 100
 111 44.0 21.4 64.3
 148 63.4 32.0 96.2



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 43.033 ug/l
 RT: 14.717 min Scan# 2170
 Delta R.T. 0.006 min
 Lab File: VN085993.D
 Acq: 14 Mar 2025 16:45

Tgt Ion: 75 Resp: 24787
 Ion Ratio Lower Upper
 75 100
 155 78.1 39.9 119.6
 157 100.6 49.9 149.7









#93
 1,2,4-Trichlorobenzene
 Concen: 44.768 ug/l
 RT: 15.388 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VN085993.D
 Acq: 14 Mar 2025 16:45




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:	180	Resp:	94648
Ion Ratio	Lower	Upper	
180	100		
182	97.7	47.3	141.9
145	32.8	16.2	48.5



#94
 Hexachlorobutadiene
 Concen: 39.523 ug/l
 RT: 15.500 min Scan# 2303
 Delta R.T. 0.006 min
 Lab File: VN085993.D
 Acq: 14 Mar 2025 16:45

Tgt Ion:	225	Resp:	50783
Ion Ratio	Lower	Upper	
225	100		
223	63.1	31.9	95.5
227	62.7	32.2	96.6









#95
Naphthalene
Concen: 46.992 ug/l
RT: 15.635 min Scan# 2326
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion:128 Resp: 272761
Ion Ratio Lower Upper
128 100
127 12.7 10.2 15.2
129 10.5 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 43.884 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. 0.000 min
Lab File: VN085993.D
Acq: 14 Mar 2025 16:45

Tgt Ion:180 Resp: 93144
Ion Ratio Lower Upper
180 100
182 95.8 47.3 141.8
145 35.1 16.9 50.6

