

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031425\
 Data File : VN085985.D
 Acq On : 14 Mar 2025 11:42
 Operator : JC\MD
 Sample : VN0314MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0314MBL01

Quant Time: Mar 14 23:59:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	130092	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	242482	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	217534	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	88652	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	105387	62.485	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	124.960%
35) Dibromofluoromethane	8.165	113	88580	55.827	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.660%
50) Toluene-d8	10.565	98	289428	50.136	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.280%
62) 4-Bromofluorobenzene	12.847	95	90675	47.673	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031425\
Data File : VN085985.D
Acq On : 14 Mar 2025 11:42
Operator : JC\MD
Sample : VN0314MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0314MBL01

Quant Time: Mar 14 23:59:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

