

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031519\
 Data File : VN054306.D
 Acq On : 15 Mar 2019 13:40
 Operator : JC/SP
 Sample : VN0315WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0315WBL01

Quant Time: Mar 16 03:33:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1383995	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2282135	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2397423	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	951324	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	703321	47.85	ug/l	0.00
Spiked Amount						
			Recovery	=	95.70%	
35) Dibromofluoromethane	7.59	113	690196	46.89	ug/l	0.00
Spiked Amount						
			Recovery	=	93.78%	
50) Toluene-d8	10.09	98	2835129	52.16	ug/l	0.00
Spiked Amount						
			Recovery	=	104.32%	
62) 4-Bromofluorobenzene	12.40	95	1065617	56.03	ug/l	0.00
Spiked Amount						
			Recovery	=	112.06%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN031519\
Data File : VN054306.D
Acq On : 15 Mar 2019 13:40
Operator : JC/SP
Sample : VN0315WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0315WBL01

Quant Time: Mar 16 03:33:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

