

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031519\
 Data File : VN054320.D
 Acq On : 15 Mar 2019 20:05
 Operator : JC/SP
 Sample : K1932-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TW-7

Quant Time: Mar 16 04:31:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1305388	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2167965	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2289382	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	940382	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	693769	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
35) Dibromofluoromethane	7.59	113	662040	47.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.70%	
50) Toluene-d8	10.09	98	2703041	52.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.68%	
62) 4-Bromofluorobenzene	12.40	95	1022271	56.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.16%	
Target Compounds						
95) Naphthalene	15.13	128	26761	4.074	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031519\
Data File : VN054320.D
Acq On : 15 Mar 2019 20:05
Operator : JC/SP
Sample : K1932-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TW-7

Quant Time: Mar 16 04:31:44 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

