

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031519\
 Data File : VN054327.D
 Acq On : 15 Mar 2019 23:18
 Operator : JC/SP
 Sample : K1930-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-4

Manual Integrations
 APPROVED

MMDadoda
 3/16/2019 11:09:28 AM

Quant Time: Mar 16 05:01:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1339402	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2288230	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2417390	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	1013004	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	733032	51.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.06%	
35) Dibromofluoromethane	7.59	113	713591	48.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.70%	
50) Toluene-d8	10.09	98	2819961	51.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.48%	
62) 4-Bromofluorobenzene	12.40	95	1081622	56.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.44%	
Target Compounds						
16) Acetone	3.82	43	33749	8.286	ug/l	99
20) Methylene Chloride	4.49	84	60680078m	4537.438	ug/l	
43) Isopropyl Acetate	8.17	43	28558	1.344	ug/l #	92
95) Naphthalene	15.14	128	22838	3.948	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031519\
Data File : VN054327.D
Acq On : 15 Mar 2019 23:18
Operator : JC/SP
Sample : K1930-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-4

Manual Integrations
APPROVED

MMDadoda
3/16/2019 11:09:28 AM

Quant Time: Mar 16 05:01:37 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

