

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031519\
 Data File : VN054328.D
 Acq On : 15 Mar 2019 23:45
 Operator : JC/SP
 Sample : K1930-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3

Quant Time: Mar 16 05:04:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1236098	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2168665	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2321462	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	953064	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	700284	53.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.70%	
35) Dibromofluoromethane	7.59	113	675140	48.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.54%	
50) Toluene-d8	10.09	98	2716122	52.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.16%	
62) 4-Bromofluorobenzene	12.40	95	1027105	56.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.66%	
Target Compounds						
16) Acetone	3.82	43	30787	8.191	ug/l	92
20) Methylene Chloride	4.54	84	5244857	424.968	ug/l	97
43) Isopropyl Acetate	8.17	43	30535	1.516	ug/l	98
95) Naphthalene	15.13	128	41497	4.394	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031519\
Data File : VN054328.D
Acq On : 15 Mar 2019 23:45
Operator : JC/SP
Sample : K1930-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: Mar 16 05:04:28 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

