

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031521\
 Data File : VN066199.D
 Acq On : 15 Mar 2021 11:49
 Operator : JC/MD
 Sample : VN0315MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0315MBL01

Quant Time: Mar 16 02:08:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 12 13:00:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.582	168	389417	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	625104	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	533601	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	203912	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.946	65	343767	54.71	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	109.42%
35) Dibromofluoromethane	7.504	113	236395	56.05	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	112.10%
50) Toluene-d8	10.025	98	883136	51.70	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.40%
62) 4-Bromofluorobenzene	12.342	95	264653	42.12	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	84.24%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031521\
Data File : VN066199.D
Acq On : 15 Mar 2021 11:49
Operator : JC/MD
Sample : VN0315MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0315MBL01

Quant Time: Mar 16 02:08:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031221W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 12 13:00:23 2021
Response via : Initial Calibration

