

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031521\
 Data File : VN066202.D
 Acq On : 15 Mar 2021 16:50
 Operator : JC/MD
 Sample : M1678-02ME
 Misc : 5.04g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 D2470-D2800(SOIL)-AME

Quant Time: Mar 16 02:10:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 12 13:00:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.579	168	299791	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	8.510	114	476329	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	422576	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	161418	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.944	65	313188	64.75	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 129.50%			
35) Dibromofluoromethane	7.504	113	205452	63.93	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 127.86%			
50) Toluene-d8	10.025	98	795526	61.12	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 122.24%			
62) 4-Bromofluorobenzene	12.345	95	240041	50.13	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 100.26%			
Target Compounds					Qvalue	
25) 2-Butanone	6.753	43	32756	8.98	ug/l	92
67) Ethyl Benzene	11.452	91	99603	6.33	ug/l	98
68) m/p-Xylenes	11.559	106	157624	27.03	ug/l	99
69) o-Xylene	11.892	106	72458	12.79	ug/l	99
86) p-Isopropyltoluene	13.233	119	23589	2.34	ug/l	97
95) Naphthalene	15.073	128	23256	4.28	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031521\
Data File : VN066202.D
Acq On : 15 Mar 2021 16:50
Operator : JC/MD
Sample : M1678-02ME
Misc : 5.04g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
D2470-D2800(SOIL)-AME

Quant Time: Mar 16 02:10:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031221W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 12 13:00:23 2021
Response via : Initial Calibration

