

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031522\
 Data File : VN071292.D
 Acq On : 15 Mar 2022 15:16
 Operator : JC\MD
 Sample : N1851-07
 Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUM-37

Quant Time: Mar 15 18:59:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.075	168	801281	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.963	114	1335554	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1165609	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	484026	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.428	65	506292	46.103	ug/l	-0.01
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.200%
35) Dibromofluoromethane	8.016	113	371019	44.023	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	88.040%
50) Toluene-d8	10.439	98	1612844	47.530	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.060%
62) 4-Bromofluorobenzene	12.727	95	536954	46.730	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031522\
Data File : VN071292.D
Acq On : 15 Mar 2022 15:16
Operator : JC\MD
Sample : N1851-07
Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUM-37

Quant Time: Mar 15 18:59:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

