

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031522\
 Data File : VN071296.D
 Acq On : 15 Mar 2022 16:52
 Operator : JC\MD
 Sample : PB143286ZHE#12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143286ZHE#12

Quant Time: Mar 16 08:02:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	767957	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1290470	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1196751	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	415617	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	492462	46.789	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	93.580%		
35) Dibromofluoromethane	8.022	113	375476	46.108	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	92.220%		
50) Toluene-d8	10.439	98	1603145	48.895	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	97.800%		
62) 4-Bromofluorobenzene	12.727	95	507851	45.742	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.480%		
Target Compounds					Qvalue	
16) Acetone	4.293	43	33056	6.455	ug/l	95
18) Methyl Acetate	4.857	43	23189	2.140	ug/l	94
20) Methylene Chloride	5.099	84	16235	2.103	ug/l	93
25) 2-Butanone	7.334	43	7787	0.944	ug/l	99
43) Isopropyl Acetate	8.563	43	129663	4.829	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031522\
Data File : VN071296.D
Acq On : 15 Mar 2022 16:52
Operator : JC\MD
Sample : PB143286ZHE#12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143286ZHE#12

Quant Time: Mar 16 08:02:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

