

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031522\
 Data File : VN071300.D
 Acq On : 15 Mar 2022 18:28
 Operator : JC\MD
 Sample : N1890-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-232

Quant Time: Mar 16 08:05:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	795714	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1346538	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1264773	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	466305	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	522081	47.873	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.740%
35) Dibromofluoromethane	8.016	113	392118	46.147	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.300%
50) Toluene-d8	10.439	98	1689129	49.372	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.740%
62) 4-Bromofluorobenzene	12.727	95	605498	52.266	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.540%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	43934	8.280	ug/l	96
18) Methyl Acetate	4.857	43	26921	2.341	ug/l #	90
20) Methylene Chloride	5.104	84	12217	1.632	ug/l #	95
43) Isopropyl Acetate	8.557	43	110281	3.936	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031522\
Data File : VN071300.D
Acq On : 15 Mar 2022 18:28
Operator : JC\MD
Sample : N1890-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-232

Quant Time: Mar 16 08:05:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

