

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031522\
 Data File : VN071302.D
 Acq On : 15 Mar 2022 19:16
 Operator : JC\MD
 Sample : N1886-03 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30969

Quant Time: Mar 16 08:06:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	791379	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1316802	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1228688	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	422152	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	509654	46.989	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.980%
35) Dibromofluoromethane	8.022	113	384789	46.307	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.620%
50) Toluene-d8	10.439	98	1610956	48.151	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.300%
62) 4-Bromofluorobenzene	12.727	95	571181	50.417	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.840%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	28698	5.438	ug/l	96
18) Methyl Acetate	4.851	43	33868	2.837	ug/l	97
51) 4-Methyl-2-Pentanone	10.328	43	31784	1.869	ug/l	98
68) m/p-Xylenes	11.945	106	7778	0.456	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031522\
Data File : VN071302.D
Acq On : 15 Mar 2022 19:16
Operator : JC\MD
Sample : N1886-03 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30969

Quant Time: Mar 16 08:06:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

