

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031522\
 Data File : VN071305.D
 Acq On : 15 Mar 2022 20:28
 Operator : JC\MD
 Sample : N1886-06 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2822-B

Quant Time: Mar 16 08:08:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	772103	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1298109	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1193809	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	431355	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	496616	46.930	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	93.860%		
35) Dibromofluoromethane	8.016	113	387124	47.259	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.520%		
50) Toluene-d8	10.439	98	1580592	47.924	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	95.840%		
62) 4-Bromofluorobenzene	12.727	95	572104	51.225	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	102.460%		
Target Compounds						Qvalue
16) Acetone	4.292	43	27249	5.292	ug/l	96
18) Methyl Acetate	4.857	43	22853	2.107	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031522\
Data File : VN071305.D
Acq On : 15 Mar 2022 20:28
Operator : JC\MD
Sample : N1886-06 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2822-B

Quant Time: Mar 16 08:08:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

