

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031522\
 Data File : VN071307.D
 Acq On : 15 Mar 2022 21:17
 Operator : JC\MD
 Sample : PB143286ZHE#13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143286ZHE#13

Quant Time: Mar 16 08:08:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	790535	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1338836	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1266804	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	464136	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	517260	47.742	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.480%		
35) Dibromofluoromethane	8.016	113	388883	46.030	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	92.060%		
50) Toluene-d8	10.439	98	1649473	48.491	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.980%		
62) 4-Bromofluorobenzene	12.727	95	559324	48.558	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	97.120%		
Target Compounds					Qvalue	
16) Acetone	4.293	43	31965	6.063	ug/l	92
18) Methyl Acetate	4.857	43	23845	2.138	ug/l #	86
20) Methylene Chloride	5.093	84	14494	1.875	ug/l	90
43) Isopropyl Acetate	8.557	43	134945	4.844	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031522\
Data File : VN071307.D
Acq On : 15 Mar 2022 21:17
Operator : JC\MD
Sample : PB143286ZHE#13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143286ZHE#13

Quant Time: Mar 16 08:08:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

