

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031522\
 Data File : VN071285.D
 Acq On : 15 Mar 2022 12:17
 Operator : JC\MD
 Sample : VN0315MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0315MBL01

Quant Time: Mar 15 18:57:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	713228	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1196796	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1084730	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	371925	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	465632	47.635	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.260%
35) Dibromofluoromethane	8.016	113	362464	47.994	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.980%
50) Toluene-d8	10.439	98	1453691	47.807	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.620%
62) 4-Bromofluorobenzene	12.727	95	460919	44.764	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.520%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031522\
Data File : VN071285.D
Acq On : 15 Mar 2022 12:17
Operator : JC\MD
Sample : VN0315MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0315MBL01

Quant Time: Mar 15 18:57:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration





Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.081 min Scan# 1036

Delta R.T. -0.005 min

Lab File: VN071285.D

Acq: 15 Mar 2022 12:17

Instrument :

MSVOA_N

ClientSampleId :





VN0315MBL01

Tgt Ion:168 Resp: 713228

Ion Ratio Lower Upper

168 100

99 57.4 41.3 61.9



1,2-Dichloroethane-d4

Concen: 47.635 ug/l

RT: 8.434 min Scan# 1096

Delta R.T. -0.006 min

Lab File: VN071285.D




Acq: 15 Mar 2022 12:17


Tgt Ion: 65 Resp: 465632

Ion Ratio Lower Upper

65 100

67 52.1 0.0 103.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.963 min Scan# 1186
Delta R.T. -0.000 min
Lab File: VN071285.D
Acq: 15 Mar 2022 12:17




Instrument :
MSVOA_N
ClientSampleId :
VN0315MBL01


Tgt Ion:114 Resp: 1196796
Ion Ratio Lower Upper
114 100
63 22.5 0.0 37.8
88 15.8 0.0 27.4



#35
Dibromofluoromethane
Concen: 47.994 ug/l
RT: 8.016 min Scan# 1025
Delta R.T. -0.006 min
Lab File: VN071285.D
Acq: 15 Mar 2022 12:17

Tgt Ion:113 Resp: 362464
Ion Ratio Lower Upper
113 100
111 103.2 82.2 123.2
192 20.2 16.9 25.3









#50
Toluene-d8
Concen: 47.807 ug/l
RT: 10.439 min Scan# 1437
Delta R.T. 0.000 min
Lab File: VN071285.D
Acq: 15 Mar 2022 12:17




Instrument :
MSVOA_N
ClientSampleId :
VN0315MBL01


Tgt Ion: 98 Resp: 1453691
Ion Ratio Lower Upper
98 100
100 65.5 50.4 75.6



#62
4-Bromofluorobenzene
Concen: 44.764 ug/l
RT: 12.727 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN071285.D
Acq: 15 Mar 2022 12:17

Tgt Ion: 95 Resp: 460919
Ion Ratio Lower Upper
95 100
174 80.2 0.0 187.2
176 76.1 0.0 181.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.739 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN071285.D
Acq: 15 Mar 2022 12:17

Instrument :
MSVOA_N
ClientSampleId :
VN0315MBL01

Tgt Ion:117 Resp: 1084730
Ion Ratio Lower Upper
117 100
82 58.2 42.4 63.6
119 31.7 25.4 38.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.669 min Scan# 1986
Delta R.T. 0.001 min
Lab File: VN071285.D
Acq: 15 Mar 2022 12:17

Tgt Ion:152 Resp: 371925
Ion Ratio Lower Upper
152 100
115 59.3 27.5 82.5
150 158.0 0.0 344.4

