

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031524\
 Data File : VN081449.D
 Acq On : 16 Mar 2024 00:52
 Operator : JC\MD
 Sample : VSTDC0050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDC0050EC

Quant Time: Mar 16 01:35:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	262157	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	502848	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	455772	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	201944	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	226608	59.783	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 119.560%	
35) Dibromofluoromethane	8.171	113	172705	56.059	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 112.120%	
50) Toluene-d8	10.571	98	594378	51.610	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.220%	
62) 4-Bromofluorobenzene	12.853	95	211353	51.924	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 103.840%	

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	2.124	85	171591	49.376	ug/l	97
3) Chloromethane	2.365	50	182686	48.667	ug/l	99
4) Vinyl Chloride	2.518	62	193643	49.450	ug/l	94
5) Bromomethane	2.947	94	116415	44.511	ug/l	97
6) Chloroethane	3.112	64	139888	52.010	ug/l	90
7) Trichlorofluoromethane	3.495	101	285390	49.264	ug/l	95
8) Diethyl Ether	3.959	74	115564	56.621	ug/l	87
9) 1,1,2-Trichlorotrifluo...	4.377	101	149684	45.373	ug/l	93
10) Methyl Iodide	4.589	142	178542	59.141	ug/l	92
11) Tert butyl alcohol	5.518	59	211014	318.368	ug/l	100
12) 1,1-Dichloroethene	4.342	96	150413	49.233	ug/l	91
13) Acrolein	4.177	56	207875	264.865	ug/l	99
14) Allyl chloride	5.024	41	238424	52.586	ug/l	89
15) Acrylonitrile	5.724	53	502947	290.466	ug/l	98
16) Acetone	4.424	43	378521	281.849	ug/l	97
17) Carbon Disulfide	4.712	76	382394	43.854	ug/l #	95
18) Methyl Acetate	5.024	43	272986	70.996	ug/l	95
19) Methyl tert-butyl Ether	5.794	73	598628	57.773	ug/l	96
20) Methylene Chloride	5.277	84	189990	57.377	ug/l	92
21) trans-1,2-Dichloroethene	5.783	96	169786	49.293	ug/l	96
22) Diisopropyl ether	6.677	45	616186	61.188	ug/l #	93
23) Vinyl Acetate	6.606	43	2382245	291.190	ug/l #	94
24) 1,1-Dichloroethane	6.565	63	347375	56.351	ug/l	98
25) 2-Butanone	7.482	43	652759	290.232	ug/l	95
26) 2,2-Dichloropropane	7.494	77	246127	45.175	ug/l	98
27) cis-1,2-Dichloroethene	7.488	96	203137	52.018	ug/l	92
28) Bromochloromethane	7.812	49	153365	58.143	ug/l #	72
29) Tetrahydrofuran	7.841	42	442415	288.471	ug/l	91
30) Chloroform	7.965	83	374488	58.540	ug/l	97
31) Cyclohexane	8.259	56	240879	41.974	ug/l #	82
32) 1,1,1-Trichloroethane	8.171	97	308625	54.580	ug/l	97
36) 1,1-Dichloropropene	8.371	75	237986	47.882	ug/l	97
37) Ethyl Acetate	7.559	43	263615	54.258	ug/l	99
38) Carbon Tetrachloride	8.359	117	252252	50.428	ug/l	98
39) Methylcyclohexane	9.606	83	230192	40.243	ug/l	96
40) Benzene	8.606	78	766229	51.648	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031524\
 Data File : VN081449.D
 Acq On : 16 Mar 2024 00:52
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Mar 16 01:35:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	153837	56.068	ug/l	98
42) 1,2-Dichloroethane	8.671	62	300229	57.396	ug/l	100
43) Isopropyl Acetate	8.694	43	454776	53.478	ug/l	97
44) Trichloroethene	9.359	130	181356	48.469	ug/l	97
45) 1,2-Dichloropropane	9.624	63	203088	54.493	ug/l	95
46) Dibromomethane	9.712	93	146954	55.942	ug/l	93
47) Bromodichloromethane	9.888	83	293246	57.503	ug/l	98
48) Methyl methacrylate	9.682	41	218973	58.254	ug/l	92
49) 1,4-Dioxane	9.700	88	80591	972.725	ug/l	89
51) 4-Methyl-2-Pentanone	10.447	43	1440308	291.691	ug/l	97
52) Toluene	10.629	92	455295	50.804	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	306314	55.515	ug/l	96
54) cis-1,3-Dichloropropene	10.312	75	320176	53.752	ug/l	95
55) 1,1,2-Trichloroethane	11.018	97	199181	56.046	ug/l	94
56) Ethyl methacrylate	10.876	69	298086	58.413	ug/l	95
57) 1,3-Dichloropropane	11.165	76	342412	55.206	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.165	63	810756	295.355	ug/l	91
59) 2-Hexanone	11.200	43	1074723	291.140	ug/l	98
60) Dibromochloromethane	11.365	129	209881	57.344	ug/l	99
61) 1,2-Dibromoethane	11.470	107	204281	55.399	ug/l	98
64) Tetrachloroethene	11.106	164	178167	49.925	ug/l	92
65) Chlorobenzene	11.894	112	485278	50.527	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.965	131	188620	53.670	ug/l	98
67) Ethyl Benzene	11.965	91	865119	50.005	ug/l	99
68) m/p-Xylenes	12.076	106	643153	98.780	ug/l	99
69) o-Xylene	12.400	106	319059	50.563	ug/l	99
70) Styrene	12.412	104	555627	55.564	ug/l	98
71) Bromoform	12.582	173	136535	55.959	ug/l #	97
73) Isopropylbenzene	12.700	105	817041	50.104	ug/l	100
74) N-amyl acetate	12.494	43	400162	58.484	ug/l	94
75) 1,1,2,2-Tetrachloroethane	12.941	83	271409	53.931	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	375333	78.688	ug/l	58
77) Bromobenzene	12.982	156	185983	48.103	ug/l	85
78) n-propylbenzene	13.041	91	978672	50.741	ug/l	98
79) 2-Chlorotoluene	13.129	91	598090	50.731	ug/l	97
80) 1,3,5-Trimethylbenzene	13.176	105	712147	52.831	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.735	75	104758	61.293	ug/l	97
82) 4-Chlorotoluene	13.223	91	590684	50.725	ug/l	96
83) tert-Butylbenzene	13.441	119	561560	48.592	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	721397	52.821	ug/l	99
85) sec-Butylbenzene	13.617	105	826656	49.445	ug/l	94
86) p-Isopropyltoluene	13.729	119	679708	50.696	ug/l	98
87) 1,3-Dichlorobenzene	13.735	146	353596	48.434	ug/l	96
88) 1,4-Dichlorobenzene	13.817	146	353732	48.476	ug/l	97
89) n-Butylbenzene	14.059	91	599406	48.750	ug/l	98
90) Hexachloroethane	14.335	117	122967	52.385	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	357883	51.093	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.723	75	59478	55.230	ug/l	87
93) 1,2,4-Trichlorobenzene	15.394	180	185156	49.453	ug/l	98
94) Hexachlorobutadiene	15.506	225	69141	44.691	ug/l	95
95) Naphthalene	15.647	128	728827	54.706	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	194719	51.284	ug/l	100


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031524\
Data File : VN081449.D
Acq On : 16 Mar 2024 00:52
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Quant Time: Mar 16 01:35:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 06 03:12:57 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.224 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion:168 Resp: 262157

Ion Ratio Lower Upper

168 100

99 65.2 59.9 89.9



#2

Dichlorodifluoromethane

Concen: 49.376 ug/l

RT: 2.124 min Scan# 29

Delta R.T. -0.000 min

Lab File: VN081449.D




Acq: 16 Mar 2024 00:52


Tgt Ion: 85 Resp: 171591

Ion Ratio Lower Upper

85 100

87 32.1 15.3 45.9









#3
Chloromethane
Concen: 48.667 ug/l
RT: 2.365 min Scan# 70
Delta R.T. 0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 50 Resp: 182686
Ion Ratio Lower Upper
50 100
52 31.8 25.0 37.4



#4
Vinyl Chloride
Concen: 49.450 ug/l
RT: 2.518 min Scan# 96
Delta R.T. 0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 62 Resp: 193643
Ion Ratio Lower Upper
62 100
64 31.8 22.7 34.1









#5
Bromomethane
Concen: 44.511 ug/l
RT: 2.947 min Scan# 169
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 94 Resp: 116415
Ion Ratio Lower Upper
94 100
96 89.2 69.2 103.8



#6
Chloroethane
Concen: 52.010 ug/l
RT: 3.112 min Scan# 197
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 64 Resp: 139888
Ion Ratio Lower Upper
64 100
66 27.3 26.4 39.6









#7
 Trichlorofluoromethane
 Concen: 49.264 ug/l
 RT: 3.495 min Scan# 20
 Delta R.T. -0.000 min
 Lab File: VN081449.D
 Acq: 16 Mar 2024 00:52




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:101 Resp: 285390
 Ion Ratio Lower Upper
 101 100
 103 63.6 54.4 81.6



#8
 Diethyl Ether
 Concen: 56.621 ug/l
 RT: 3.959 min Scan# 341
 Delta R.T. 0.006 min
 Lab File: VN081449.D
 Acq: 16 Mar 2024 00:52

Tgt Ion: 74 Resp: 115564
 Ion Ratio Lower Upper
 74 100
 45 93.5 53.3 159.9









#9
1,1,2-Trichlorotrifluoroethane
Concen: 45.373 ug/l
RT: 4.377 min Scan# 410
Delta R.T. 0.012 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 149684
Ion Ratio Lower Upper
101 100
85 46.4 38.1 57.1
151 69.7 49.1 73.7



#10
Methyl Iodide
Concen: 59.141 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion:142 Resp: 178542
Ion Ratio Lower Upper
142 100
127 52.9 37.0 55.4
141 17.0 12.5 18.7









#11
Tert butyl alcohol
Concen: 318.368 ug/l
RT: 5.518 min Scan# 605
Delta R.T. 0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 59 Resp: 211014
Ion Ratio Lower Upper
59 100
57 10.9 8.6 12.8



#12
1,1-Dichloroethene
Concen: 49.233 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 96 Resp: 150413
Ion Ratio Lower Upper
96 100
61 156.0 138.2 207.4
98 60.8 49.2 73.8









#13
Acrolein
Concen: 264.865 ug/l
RT: 4.177 min Scan# 31
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 207875
Ion Ratio Lower Upper
56 100
55 69.8 55.3 82.9



#14
Allyl chloride
Concen: 52.586 ug/l
RT: 5.024 min Scan# 522
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 41 Resp: 238424
Ion Ratio Lower Upper
41 100
39 81.2 57.7 86.5
76 39.7 26.8 40.2









#15
Acrylonitrile
Concen: 290.466 ug/l
RT: 5.724 min Scan# 641
Delta R.T. 0.012 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 53 Resp: 502947
Ion Ratio Lower Upper
53 100
52 82.1 64.8 97.2
51 36.2 26.9 40.3



#16
Acetone
Concen: 281.849 ug/l
RT: 4.424 min Scan# 420
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 43 Resp: 378521
Ion Ratio Lower Upper
43 100
58 29.6 24.8 37.2









#17
Carbon Disulfide
Concen: 43.854 ug/l
RT: 4.712 min Scan# 469
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 76 Resp: 382394
Ion Ratio Lower Upper
76 100
78 9.5 6.3 9.5#



#18
Methyl Acetate
Concen: 70.996 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 43 Resp: 272986
Ion Ratio Lower Upper
43 100
74 23.8 17.3 25.9









#19
Methyl tert-butyl Ether
Concen: 57.773 ug/l
RT: 5.794 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 73 Resp: 598628
Ion Ratio Lower Upper
73 100
57 20.7 18.2 27.4



#20
Methylene Chloride
Concen: 57.377 ug/l
RT: 5.277 min Scan# 565
Delta R.T. -0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 84 Resp: 189990
Ion Ratio Lower Upper
84 100
49 121.9 108.1 162.1
51 37.2 32.9 49.3
86 63.6 48.6 73.0









#21
trans-1,2-Dichloroethene
Concen: 49.293 ug/l
RT: 5.783 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 96 Resp: 169786
Ion Ratio Lower Upper
96 100
61 133.5 112.8 169.2
98 63.6 51.0 76.4



#22
Diisopropyl ether
Concen: 61.188 ug/l
RT: 6.677 min Scan# 803
Delta R.T. 0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 45 Resp: 616186
Ion Ratio Lower Upper
45 100
43 56.6 41.8 62.8
87 28.1 18.1 27.1#
59 12.2 9.0 13.6









#23
 Vinyl Acetate
 Concen: 291.190 ug/l
 RT: 6.606 min Scan# 791
 Delta R.T. 0.006 min
 Lab File: VN081449.D
 Acq: 16 Mar 2024 00:52




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 43 Resp: 2382245
 Ion Ratio Lower Upper
 43 100
 86 11.1 7.1 10.7#



#24
 1,1-Dichloroethane
 Concen: 56.351 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. -0.000 min
 Lab File: VN081449.D
 Acq: 16 Mar 2024 00:52

Tgt Ion: 63 Resp: 347375
 Ion Ratio Lower Upper
 63 100
 98 6.5 2.9 8.6
 100 4.2 1.9 5.7









#25
2-Butanone
Concen: 290.232 ug/l
RT: 7.482 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 652759
Ion Ratio Lower Upper
43 100
72 26.8 19.5 29.3



#26
2,2-Dichloropropane
Concen: 45.175 ug/l
RT: 7.494 min Scan# 942
Delta R.T. 0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 77 Resp: 246127
Ion Ratio Lower Upper
77 100
97 22.2 11.5 34.5









#27
 cis-1,2-Dichloroethene
 Concen: 52.018 ug/l
 RT: 7.488 min Scan# 941
 Delta R.T. -0.000 min
 Lab File: VN081449.D
 Acq: 16 Mar 2024 00:52




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 96 Resp: 203137
 Ion Ratio Lower Upper
 96 100
 61 149.0 0.0 325.4
 98 65.9 0.0 130.4



#28
 Bromochloromethane
 Concen: 58.143 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. -0.006 min
 Lab File: VN081449.D
 Acq: 16 Mar 2024 00:52

Tgt Ion: 49 Resp: 153365
 Ion Ratio Lower Upper
 49 100
 129 1.7 0.0 1.6#
 130 72.9 42.3 63.5#





#29

Tetrahydrofuran

Concen: 288.471 ug/l

RT: 7.841 min Scan# 1001

Delta R.T. -0.000 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 42 Resp: 442415

Ion Ratio Lower Upper

42 100

72 44.2 31.1 46.7

71 41.5 28.3 42.5



#30

Chloroform

Concen: 58.540 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. -0.006 min

Lab File: VN081449.D




Acq: 16 Mar 2024 00:52


Tgt Ion: 83 Resp: 374488

Ion Ratio Lower Upper

83 100

85 61.1 51.0 76.6









#31
Cyclohexane
Concen: 41.974 ug/l
RT: 8.259 min Scan# 1072
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 240879
Ion Ratio Lower Upper
56 100
69 33.4 20.1 30.1#
84 85.8 56.9 85.3#



#32
1,1,1-Trichloroethane
Concen: 54.580 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 97 Resp: 308625
Ion Ratio Lower Upper
97 100
99 65.9 53.4 80.0
61 50.7 43.8 65.8









#33
 1,2-Dichloroethane-d4
 Concen: 59.783 ug/l
 RT: 8.582 min Scan# 1126
 Delta R.T. 0.006 min
 Lab File: VN081449.D
 Acq: 16 Mar 2024 00:52




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 65 Resp: 226608
 Ion Ratio Lower Upper
 65 100
 67 50.8 0.0 102.4



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 9.106 min Scan# 1216
 Delta R.T. -0.000 min
 Lab File: VN081449.D
 Acq: 16 Mar 2024 00:52

Tgt Ion: 114 Resp: 502848
 Ion Ratio Lower Upper
 114 100
 63 22.4 0.0 48.0
 88 15.0 0.0 34.8









#35
Dibromofluoromethane
Concen: 56.059 ug/l
RT: 8.171 min Scan# 1056
Delta R.T. 0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:113 Resp: 172705
Ion Ratio Lower Upper
113 100
111 102.2 82.2 123.4
192 15.9 12.5 18.7



#36
1,1-Dichloropropene
Concen: 47.882 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. -0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 75 Resp: 237986
Ion Ratio Lower Upper
75 100
110 33.0 16.0 48.0
77 30.1 25.8 38.6









#37
Ethyl Acetate
Concen: 54.258 ug/l
RT: 7.559 min Scan# 91
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 263615
Ion Ratio Lower Upper
43 100
61 14.1 10.6 15.8
70 9.9 8.0 12.0



#38
Carbon Tetrachloride
Concen: 50.428 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. -0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 117 Resp: 252252
Ion Ratio Lower Upper
117 100
119 93.6 73.4 110.2
121 30.9 24.4 36.6





#39

Methylcyclohexane

Concen: 40.243 ug/l

RT: 9.606 min Scan# 1130

Delta R.T. 0.006 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 83 Resp: 230192

Ion Ratio Lower Upper

83 100

55 81.1 65.7 98.5

98 47.4 33.8 50.6



#40

Benzene

Concen: 51.648 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN081449.D




Acq: 16 Mar 2024 00:52


Tgt Ion: 78 Resp: 766229

Ion Ratio Lower Upper

78 100

77 24.5 20.4 30.6









#41
Methacrylonitrile
Concen: 56.068 ug/l
RT: 7.777 min Scan# 990
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 41 Resp: 153837
Ion Ratio Lower Upper
41 100
39 56.2 48.4 72.6
67 71.2 56.8 85.2
52 30.9 25.0 37.4



#42
1,2-Dichloroethane
Concen: 57.396 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 62 Resp: 300229
Ion Ratio Lower Upper
62 100
98 9.0 0.0 18.0









#43
Isopropyl Acetate
Concen: 53.478 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. 0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 454776
Ion Ratio Lower Upper
43 100
61 25.6 19.9 29.9
87 13.3 9.0 13.6



#44
Trichloroethene
Concen: 48.469 ug/l
RT: 9.359 min Scan# 1259
Delta R.T. 0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion:130 Resp: 181356
Ion Ratio Lower Upper
130 100
95 102.3 0.0 210.4





#45

1,2-Dichloropropane

Concen: 54.493 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. -0.000 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 63 Resp: 203088

Ion Ratio Lower Upper

63 100

65 29.8 26.1 39.1



#46

Dibromomethane

Concen: 55.942 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.000 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52




Tgt Ion: 93 Resp: 146954


Ion Ratio Lower Upper

93 100

95 80.3 68.1 102.1

174 79.1 57.2 85.8





#47

Bromodichloromethane

Concen: 57.503 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. -0.000 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 83 Resp: 293246

Ion Ratio Lower Upper

83 100

85 64.1 52.3 78.5

127 6.9 4.6 7.0



#48

Methyl methacrylate

Concen: 58.254 ug/l

RT: 9.682 min Scan# 1314

Delta R.T. -0.000 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52




Tgt Ion: 41 Resp: 218973


Ion Ratio Lower Upper

41 100

69 83.3 62.2 93.2

39 66.8 47.6 71.4





#49

1,4-Dioxane

Concen: 972.725 ug/l

RT: 9.700 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 88 Resp: 80591

Ion Ratio Lower Upper

88 100

43 30.5 30.1 45.1

58 70.0 63.9 95.9



#50

Toluene-d8

Concen: 51.610 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN081449.D




Acq: 16 Mar 2024 00:52


Tgt Ion: 98 Resp: 594378

Ion Ratio Lower Upper

98 100

100 65.8 51.4 77.0









#51
4-Methyl-2-Pentanone
Concen: 291.691 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 1440308
Ion Ratio Lower Upper
43 100
58 38.4 29.4 44.0



#52
Toluene
Concen: 50.804 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 92 Resp: 455295
Ion Ratio Lower Upper
92 100
91 175.6 139.4 209.0









#53
 t-1,3-Dichloropropene
 Concen: 55.515 ug/l
 RT: 10.835 min Scan# 1510
 Delta R.T. -0.000 min
 Lab File: VN081449.D
 Acq: 16 Mar 2024 00:52




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 75 Resp: 306314
 Ion Ratio Lower Upper
 75 100
 77 29.4 25.5 38.3



#54
 cis-1,3-Dichloropropene
 Concen: 53.752 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. -0.000 min
 Lab File: VN081449.D
 Acq: 16 Mar 2024 00:52

Tgt Ion: 75 Resp: 320176
 Ion Ratio Lower Upper
 75 100
 77 30.4 23.3 34.9
 39 51.3 44.6 66.8





#55

1,1,2-Trichloroethane

Concen: 56.046 ug/l

RT: 11.018 min Scan# 1541

Delta R.T. -0.000 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 97 Resp: 199181





Ion Ratio Lower Upper

97 100

83 84.9 72.9 109.3

85 55.1 50.2 75.2

99 61.0 49.6 74.4



#56

Ethyl methacrylate

Concen: 58.413 ug/l

RT: 10.876 min Scan# 1517

Delta R.T. -0.000 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52




Tgt Ion: 69 Resp: 298086


Ion Ratio Lower Upper

69 100

41 71.5 61.0 91.6

39 42.6 31.8 47.6





#57

1,3-Dichloropropane

Concen: 55.206 ug/l

RT: 11.165 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 76 Resp: 342412

Ion Ratio Lower Upper

76 100

78 32.1 25.5 38.3



#58

2-Chloroethyl Vinyl ether

Concen: 295.355 ug/l

RT: 10.165 min Scan# 1396

Delta R.T. 0.006 min

Lab File: VN081449.D




Acq: 16 Mar 2024 00:52


Tgt Ion: 63 Resp: 810756

Ion Ratio Lower Upper

63 100

106 27.8 18.6 28.0









#59
2-Hexanone
Concen: 291.140 ug/l
RT: 11.200 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 1074723
Ion Ratio Lower Upper
43 100
58 53.1 25.9 77.8



#60
Dibromochloromethane
Concen: 57.344 ug/l
RT: 11.365 min Scan# 1600
Delta R.T. 0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 129 Resp: 209881
Ion Ratio Lower Upper
129 100
127 79.4 39.4 118.2









#61
1,2-Dibromoethane
Concen: 55.399 ug/l
RT: 11.470 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:107 Resp: 204281
Ion Ratio Lower Upper
107 100
109 93.4 73.4 110.2



#62
4-Bromofluorobenzene
Concen: 51.924 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 95 Resp: 211353
Ion Ratio Lower Upper
95 100
174 71.4 0.0 120.4
176 69.0 0.0 121.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.870 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:117 Resp: 455772

Ion Ratio Lower Upper

117 100

82 54.3 52.7 79.1

119 30.4 25.3 37.9



#64

Tetrachloroethene

Concen: 49.925 ug/l

RT: 11.106 min Scan# 1556

Delta R.T. -0.000 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52

Tgt Ion:164 Resp: 178167




Ion Ratio Lower Upper


164 100

166 122.8 105.8 158.6

129 105.1 89.6 134.4

131 96.6 84.9 127.3









#65
Chlorobenzene
Concen: 50.527 ug/l
RT: 11.894 min Scan# 1690
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:112 Resp: 485278
Ion Ratio Lower Upper
112 100
114 31.7 23.7 35.5



#66
1,1,1,2-Tetrachloroethane
Concen: 53.670 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion:131 Resp: 188620
Ion Ratio Lower Upper
131 100
133 94.0 46.8 140.3
119 66.0 34.6 103.9





#67

Ethyl Benzene

Concen: 50.005 ug/l

RT: 11.965 min Scan# 1702

Delta R.T. -0.000 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 91 Resp: 865119

Ion Ratio Lower Upper

91 100

106 29.2 23.8 35.6



#68

m/p-Xylenes

Concen: 98.780 ug/l

RT: 12.076 min Scan# 1721

Delta R.T. 0.006 min

Lab File: VN081449.D




Acq: 16 Mar 2024 00:52


Tgt Ion:106 Resp: 643153

Ion Ratio Lower Upper

106 100

91 210.3 169.0 253.6









#69
o-Xylene
Concen: 50.563 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:106 Resp: 319059
Ion Ratio Lower Upper
106 100
91 222.2 112.1 336.3



#70
Styrene
Concen: 55.564 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion:104 Resp: 555627
Ion Ratio Lower Upper
104 100
78 52.0 43.8 65.8
103 54.8 44.2 66.2





#71

Bromoform

Concen: 55.959 ug/l

RT: 12.582 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:173 Resp: 136535

Ion Ratio Lower Upper

173 100

175 48.1 23.1 69.2

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. -0.000 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52




Tgt Ion:152 Resp: 201944


Ion Ratio Lower Upper

152 100

115 65.2 32.3 96.8

150 177.8 0.0 339.8









#73
Isopropylbenzene
Concen: 50.104 ug/l
RT: 12.700 min Scan# 1827
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:105 Resp: 817041
Ion Ratio Lower Upper
105 100
120 25.2 12.6 37.6



#74
N-ethyl acetate
Concen: 58.484 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 43 Resp: 400162
Ion Ratio Lower Upper
43 100
70 41.1 30.1 45.1
55 25.2 17.0 25.4
61 23.3 17.5 26.3









#75
 1,1,2,2-Tetrachloroethane
 Concen: 53.931 ug/l
 RT: 12.941 min Scan# 1868
 Delta R.T. -0.000 min
 Lab File: VN081449.D
 Acq: 16 Mar 2024 00:52




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 83 Resp: 271409
 Ion Ratio Lower Upper
 83 100
 131 9.8 4.1 12.3
 85 65.0 32.1 96.3



#76
 1,2,3-Trichloropropane
 Concen: 78.688 ug/l
 RT: 12.994 min Scan# 1877
 Delta R.T. -0.006 min
 Lab File: VN081449.D
 Acq: 16 Mar 2024 00:52

Tgt Ion: 75 Resp: 375333
 Ion Ratio Lower Upper
 75 100
 77 121.4 91.0 272.8









#77
Bromobenzene
Concen: 48.103 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:156 Resp: 185983
Ion Ratio Lower Upper
156 100
77 244.9 138.1 414.3
158 100.7 46.5 139.5



#78
n-propylbenzene
Concen: 50.741 ug/l
RT: 13.041 min Scan# 1885
Delta R.T. 0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 91 Resp: 978672
Ion Ratio Lower Upper
91 100
120 21.5 10.3 30.9









#79
2-Chlorotoluene
Concen: 50.731 ug/l
RT: 13.129 min Scan# 1900
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 598090
Ion Ratio Lower Upper
91 100
126 31.0 14.8 44.4



#80
1,3,5-Trimethylbenzene
Concen: 52.831 ug/l
RT: 13.176 min Scan# 1908
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 105 Resp: 712147
Ion Ratio Lower Upper
105 100
120 47.8 23.0 69.0









#81
trans-1,4-Dichloro-2-butene
Concen: 61.293 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 75 Resp: 104758
Ion Ratio Lower Upper
75 100
53 92.1 71.4 107.2
89 39.5 33.4 50.0



#82
4-Chlorotoluene
Concen: 50.725 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 91 Resp: 590684
Ion Ratio Lower Upper
91 100
126 32.8 15.3 45.9





#83

tert-Butylbenzene

Concen: 48.592 ug/l

RT: 13.441 min Scan# 1953

Delta R.T. -0.000 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:119 Resp: 561560

Ion Ratio Lower Upper

119 100

91 72.9 36.8 110.4

134 26.3 12.4 37.4



#84

1,2,4-Trimethylbenzene

Concen: 52.821 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. -0.000 min

Lab File: VN081449.D




Acq: 16 Mar 2024 00:52


Tgt Ion:105 Resp: 721397

Ion Ratio Lower Upper

105 100

120 42.6 20.9 62.7





#85

sec-Butylbenzene

Concen: 49.445 ug/l

RT: 13.617 min Scan# 1983

Delta R.T. -0.000 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion:105 Resp: 826656

Ion Ratio Lower Upper

105 100

134 19.8 8.6 25.8



#86

p-Isopropyltoluene

Concen: 50.696 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52




Tgt Ion:119 Resp: 679708


Ion Ratio Lower Upper

119 100

134 25.7 12.4 37.2

91 25.6 13.5 40.4









#87
 1,3-Dichlorobenzene
 Concen: 48.434 ug/l
 RT: 13.735 min Scan# 2003
 Delta R.T. -0.000 min
 Lab File: VN081449.D
 Acq: 16 Mar 2024 00:52




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:146 Resp: 353596
 Ion Ratio Lower Upper
 146 100
 111 43.6 23.9 71.7
 148 63.5 32.6 97.7



#88
 1,4-Dichlorobenzene
 Concen: 48.476 ug/l
 RT: 13.817 min Scan# 2017
 Delta R.T. 0.006 min
 Lab File: VN081449.D
 Acq: 16 Mar 2024 00:52

Tgt Ion:146 Resp: 353732
 Ion Ratio Lower Upper
 146 100
 111 42.5 22.6 67.7
 148 62.1 31.9 95.5









#89
n-Butylbenzene
Concen: 48.750 ug/l
RT: 14.059 min Scan# 2058
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91	Resp: 599406
Ion Ratio	Lower Upper
91	100
92	52.3 26.7 80.1
134	23.6 10.8 32.4



#90
Hexachloroethane
Concen: 52.385 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 117	Resp: 122967
Ion Ratio	Lower Upper
117	100
201	61.1 30.0 90.0









#91
1,2-Dichlorobenzene
Concen: 51.093 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:146 Resp: 357883
Ion Ratio Lower Upper
146 100
111 44.9 23.9 71.8
148 63.0 31.4 94.0



#92
1,2-Dibromo-3-Chloropropane
Concen: 55.230 ug/l
RT: 14.723 min Scan# 2171
Delta R.T. 0.006 min
Lab File: VN081449.D
Acq: 16 Mar 2024 00:52

Tgt Ion: 75 Resp: 59478
Ion Ratio Lower Upper
75 100
155 74.2 31.8 95.4
157 93.3 41.1 123.4









#93
 1,2,4-Trichlorobenzene
 Concen: 49.453 ug/l
 RT: 15.394 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: VN081449.D
 Acq: 16 Mar 2024 00:52




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:	180	Resp:	185156
Ion Ratio	Lower	Upper	
180	100		
182	95.1	46.6	139.7
145	33.2	16.8	50.3



#94
 Hexachlorobutadiene
 Concen: 44.691 ug/l
 RT: 15.506 min Scan# 2304
 Delta R.T. 0.006 min
 Lab File: VN081449.D
 Acq: 16 Mar 2024 00:52

Tgt Ion:	225	Resp:	69141
Ion Ratio	Lower	Upper	
225	100		
223	63.2	33.3	99.8
227	61.8	33.1	99.2





#95

Naphthalene

Concen: 54.706 ug/l

RT: 15.647 min Scan# 2327

Delta R.T. 0.006 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:128 Resp: 728827

Ion Ratio Lower Upper

128 100

127 13.1 10.0 15.0

129 11.1 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 51.284 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. -0.000 min

Lab File: VN081449.D

Acq: 16 Mar 2024 00:52

Tgt Ion:180 Resp: 194719

Ion Ratio Lower Upper

180 100

182 93.3 46.5 139.3

145 35.3 17.8 53.4

