

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN031524\  
 Data File : VN081439.D  
 Acq On : 15 Mar 2024 20:53  
 Operator : JC\MD  
 Sample : P1772-11 50X  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 1349

Quant Time: Mar 15 23:39:58 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N030524W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Mar 06 03:12:57 2024  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 8.230  | 168  | 310953   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 9.106  | 114  | 566429   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 11.870 | 117  | 488659   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 13.794 | 152  | 188493   | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 33) 1,2-Dichloroethane-d4   | 8.582  | 65    | 236592   | 52.622   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =    | 105.240% |
| 35) Dibromofluoromethane    | 8.171  | 113   | 181769   | 52.378   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =    | 104.760% |
| 50) Toluene-d8              | 10.571 | 98    | 690319   | 53.212   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =    | 106.420% |
| 62) 4-Bromofluorobenzene    | 12.853 | 95    | 211884   | 46.211   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 64 - 133 | Recovery | =    | 92.420%  |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN031524\  
Data File : VN081439.D  
Acq On : 15 Mar 2024 20:53  
Operator : JC\MD  
Sample : P1772-11 50X  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 16 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
1349

Quant Time: Mar 15 23:39:58 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N030524W.M  
Quant Title : SW846 8260  
QLast Update : Wed Mar 06 03:12:57 2024  
Response via : Initial Calibration

