

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031524\
 Data File : VN081446.D
 Acq On : 15 Mar 2024 23:40
 Operator : JC\MD
 Sample : P1773-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4B

Quant Time: Mar 16 01:33:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	264585	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	485536	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	420566	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	149345	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	213521	55.814	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.620%
35) Dibromofluoromethane	8.171	113	161879	54.418	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.840%
50) Toluene-d8	10.571	98	584747	52.584	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.160%
62) 4-Bromofluorobenzene	12.853	95	168865	42.965	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	85.920%
Target Compounds						
						Qvalue
16) Acetone	4.442	43	4406	3.251	ug/l	91
30) Chloroform	7.971	83	13342	2.066	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031524\
Data File : VN081446.D
Acq On : 15 Mar 2024 23:40
Operator : JC\MD
Sample : P1773-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-4B

Quant Time: Mar 16 01:33:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 06 03:12:57 2024
Response via : Initial Calibration

