

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031619\
 Data File : VN054336.D
 Acq On : 16 Mar 2019 14:56
 Operator : JC/SP
 Sample : K1962-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Mar 18 05:43:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1813298	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2650941	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2175796	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	753357	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	899169	46.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.38%	
35) Dibromofluoromethane	7.59	113	816105	47.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.46%	
50) Toluene-d8	10.09	98	3037532	48.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.22%	
62) 4-Bromofluorobenzene	12.40	95	882667	39.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	79.90%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031619\
Data File : VN054336.D
Acq On : 16 Mar 2019 14:56
Operator : JC/SP
Sample : K1962-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Mar 18 05:43:05 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

