

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031619\
 Data File : VN054350.D
 Acq On : 16 Mar 2019 21:19
 Operator : JC/SP
 Sample : K1899-03DL 50X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4ADL

Quant Time: Mar 18 12:16:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1868754	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2685385	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2236085	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	882593	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	903687	45.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.06%	
35) Dibromofluoromethane	7.59	113	811963	46.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.76%	
50) Toluene-d8	10.09	98	3088707	48.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.58%	
62) 4-Bromofluorobenzene	12.40	95	939153	41.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.92%	
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.04	105	147478	2.559	ug/l	100
95) Naphthalene	15.13	128	1743862	45.392	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031619\
Data File : VN054350.D
Acq On : 16 Mar 2019 21:19
Operator : JC/SP
Sample : K1899-03DL 50X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-4ADL

Quant Time: Mar 18 12:16:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

