

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031619\
 Data File : VN054356.D
 Acq On : 17 Mar 2019 00:58
 Operator : JC/SP
 Sample : VN0316WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0316WBL02

Quant Time: Mar 18 07:59:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1871214	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2668496	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2209833	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	779547	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	908335	45.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.42%	
35) Dibromofluoromethane	7.59	113	819399	47.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.22%	
50) Toluene-d8	10.09	98	3053735	48.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.08%	
62) 4-Bromofluorobenzene	12.40	95	894233	40.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	80.42%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031619\
Data File : VN054356.D
Acq On : 17 Mar 2019 00:58
Operator : JC/SP
Sample : VN0316WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0316WBL02

Quant Time: Mar 18 07:59:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

