

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031619\
 Data File : VN054364.D
 Acq On : 17 Mar 2019 4:38
 Operator : JC/SP
 Sample : K1956-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-8

Quant Time: Mar 18 08:20:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1848741	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2670687	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2267108	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	849192	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	913474	46.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.06%	
35) Dibromofluoromethane	7.59	113	813277	47.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.44%	
50) Toluene-d8	10.09	98	3061052	48.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.24%	
62) 4-Bromofluorobenzene	12.40	95	935324	42.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.04%	
Target Compounds						
16) Acetone	3.82	43	20017	3.561	ug/l	88
20) Methylene Chloride	4.55	84	15044881	815.057	ug/l	99
43) Isopropyl Acetate	8.17	43	27646	1.115	ug/l	98
95) Naphthalene	15.13	128	23786	4.064	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031619\
Data File : VN054364.D
Acq On : 17 Mar 2019 4:38
Operator : JC/SP
Sample : K1956-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-8

Quant Time: Mar 18 08:20:54 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

