

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031619\
 Data File : VN054367.D
 Acq On : 17 Mar 2019 6:00
 Operator : JC/SP
 Sample : K1958-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Mar 18 08:30:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1819485	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2646319	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2238771	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	834795	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	897850	46.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.94%	
35) Dibromofluoromethane	7.59	113	792934	46.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.92%	
50) Toluene-d8	10.09	98	3050900	48.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.80%	
62) 4-Bromofluorobenzene	12.40	95	912099	41.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.70%	
Target Compounds						
20) Methylene Chloride	4.54	84	97104	5.345	ug/l	96
43) Isopropyl Acetate	8.17	43	25899	1.054	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031619\
Data File : VN054367.D
Acq On : 17 Mar 2019 6:00
Operator : JC/SP
Sample : K1958-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Mar 18 08:30:24 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

