

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031620\
 Data File : VN060525.D
 Acq On : 16 Mar 2020 16:00
 Operator : JC/MD
 Sample : L1942-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-10-DA

Quant Time: Mar 17 09:27:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	211996	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	348365	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	326001	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	148208	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	136710	50.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.52%	
35) Dibromofluoromethane	7.57	113	102770	63.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	127.12%	
50) Toluene-d8	10.08	98	430433	52.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.20%	
62) 4-Bromofluorobenzene	12.40	95	160935	51.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.76%	
Target Compounds						
16) Acetone	3.78	43	12781	14.551	ug/l	95
20) Methylene Chloride	4.53	84	8245	3.418	ug/l	97
43) Isopropyl Acetate	8.14	43	61295	11.137	ug/l #	90
51) 4-Methyl-2-Pentanone	9.97	43	5152	1.651	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031620\
Data File : VN060525.D
Acq On : 16 Mar 2020 16:00
Operator : JC/MD
Sample : L1942-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-10-DA

Quant Time: Mar 17 09:27:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

