

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031620\
 Data File : VN060526.D
 Acq On : 16 Mar 2020 16:22
 Operator : JC/MD
 Sample : L1942-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3-SA

Quant Time: Mar 17 09:29:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	216020	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	358689	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	335267	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	151345	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	140100	50.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.10%	
35) Dibromofluoromethane	7.56	113	106286	63.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	127.68%	
50) Toluene-d8	10.08	98	441923	52.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.90%	
62) 4-Bromofluorobenzene	12.40	95	163984	51.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.68%	
Target Compounds						
16) Acetone	3.78	43	21990	24.569	ug/l	98
20) Methylene Chloride	4.52	84	8629	3.511	ug/l	91
43) Isopropyl Acetate	8.14	43	63074	11.131	ug/l #	90
51) 4-Methyl-2-Pentanone	9.97	43	5202	1.619	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031620\
Data File : VN060526.D
Acq On : 16 Mar 2020 16:22
Operator : JC/MD
Sample : L1942-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3-SA

Quant Time: Mar 17 09:29:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

