

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031620\
 Data File : VN060528.D
 Acq On : 16 Mar 2020 17:07
 Operator : JC/MD
 Sample : PB127541ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127541ZHE#01

Quant Time: Mar 17 09:32:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	211373	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	352558	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	330153	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	150155	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	136331	50.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.54%	
35) Dibromofluoromethane	7.56	113	102016	62.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	124.68%	
50) Toluene-d8	10.08	98	435414	52.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.14%	
62) 4-Bromofluorobenzene	12.40	95	162785	51.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.70%	
Target Compounds						
16) Acetone	3.79	43	7207	8.229	ug/l	96
20) Methylene Chloride	4.52	84	4638	1.928	ug/l	92
43) Isopropyl Acetate	8.14	43	69859	12.542	ug/l #	91
51) 4-Methyl-2-Pentanone	9.97	43	6255	1.981	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031620\
Data File : VN060528.D
Acq On : 16 Mar 2020 17:07
Operator : JC/MD
Sample : PB127541ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127541ZHE#01

Quant Time: Mar 17 09:32:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

