

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031620\
 Data File : VN060529.D
 Acq On : 16 Mar 2020 17:30
 Operator : JC/MD
 Sample : PB127541ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127541ZHE#02

Quant Time: Mar 17 09:34:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	214655	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	360995	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	338107	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	154476	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	139489	50.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.30%	
35) Dibromofluoromethane	7.56	113	105329	62.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	125.72%	
50) Toluene-d8	10.08	98	446882	52.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.38%	
62) 4-Bromofluorobenzene	12.40	95	167790	52.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.38%	
Target Compounds						
16) Acetone	3.79	43	6984	7.853	ug/l	98
20) Methylene Chloride	4.51	84	4145	1.697	ug/l	91
43) Isopropyl Acetate	8.14	43	71244	12.492	ug/l #	90
51) 4-Methyl-2-Pentanone	9.97	43	6283	1.943	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031620\
Data File : VN060529.D
Acq On : 16 Mar 2020 17:30
Operator : JC/MD
Sample : PB127541ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127541ZHE#02

Quant Time: Mar 17 09:34:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

