

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031620\
 Data File : VN060534.D
 Acq On : 16 Mar 2020 19:23
 Operator : JC/MD
 Sample : PB127541ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127541ZHE#07

Quant Time: Mar 17 09:27:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	200644	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	334256	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	310651	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	137872	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	130554	50.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.42%	
35) Dibromofluoromethane	7.57	113	98859	63.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	127.44%	
50) Toluene-d8	10.08	98	415300	52.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.78%	
62) 4-Bromofluorobenzene	12.40	95	149781	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
Target Compounds						
16) Acetone	3.79	43	5491	6.605	ug/l	95
20) Methylene Chloride	4.53	84	4060	1.778	ug/l	98
43) Isopropyl Acetate	8.14	43	49748	9.421	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031620\
Data File : VN060534.D
Acq On : 16 Mar 2020 19:23
Operator : JC/MD
Sample : PB127541ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127541ZHE#07

Quant Time: Mar 17 09:27:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

