

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031621\
 Data File : VN066212.D
 Acq On : 16 Mar 2021 13:18
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Mar 17 05:21:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.590	168	118699	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	202013	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	180210	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	59042	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	85991	50.28	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.56%
35) Dibromofluoromethane	7.512	113	64328	49.92	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.84%
50) Toluene-d8	10.028	98	251119	50.39	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.78%
62) 4-Bromofluorobenzene	12.345	95	75257	41.90	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	83.80%

Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.821	85	566	0.48	ug/l	84
3) Chloromethane	2.017	50	5056	2.39	ug/l	99
4) Vinyl Chloride	2.145	62	1210	0.72	ug/l	91
5) Bromomethane	2.529	94	2238	2.32	ug/l #	75
7) Trichlorofluoromethane	2.969	101	861	0.44	ug/l #	77
10) Methyl Iodide	3.878	142	1486	1.00	ug/l #	47
11) Tert butyl alcohol	4.723	59	83	0.28	ug/l #	73
13) Acrolein	3.532	56	3335	Below Cal	#	70
14) Allyl chloride	4.229	41	614	0.23	ug/l #	36
15) Acrylonitrile	4.886	53	2957	3.38	ug/l #	49
16) Acetone	3.738	43	5110	6.22	ug/l	100
17) Carbon Disulfide	3.969	76	12058	3.83	ug/l	98
20) Methylene Chloride	4.452	84	1280	1.00	ug/l #	66
21) trans-1,2-Dichloroethene	4.935	96	1917	1.55	ug/l #	74
25) 2-Butanone	6.753	43	4268	3.34	ug/l #	74
27) cis-1,2-Dichloroethene	6.734	96	1100	0.77	ug/l	45
28) Bromochloromethane	7.107	49	584	0.41	ug/l #	47
29) Tetrahydrofuran	7.131	42	835	0.95	ug/l #	57
30) Chloroform	7.292	83	936	0.38	ug/l #	82
31) Cyclohexane	7.584	56	3195	1.25	ug/l #	1
36) 1,1-Dichloropropene	7.713	75	1601	0.78	ug/l #	67
38) Carbon Tetrachloride	7.584	117	10873	5.93	ug/l #	18
39) Methylcyclohexane	9.008	83	1152	0.52	ug/l #	81
42) 1,2-Dichloroethane	8.051	62	1223	0.55	ug/l #	79
43) Isopropyl Acetate	8.091	43	1473	0.33	ug/l #	77
44) Trichloroethene	8.759	130	794	0.53	ug/l	67
46) Dibromomethane	9.135	93	337	0.34	ug/l #	1
47) Bromodichloromethane	9.338	83	542	0.25	ug/l #	76
48) Methyl methacrylate	9.137	41	1392	0.65	ug/l #	61
49) 1,4-Dioxane	9.140	88	292	12.56	ug/l #	16
51) 4-Methyl-2-Pentanone	9.920	43	10322	3.78	ug/l	95
52) Toluene	10.089	92	1250	0.35	ug/l #	44
53) t-1,3-Dichloropropene	10.320	75	1954	0.80	ug/l #	83
54) cis-1,3-Dichloropropene	9.773	75	1197	0.45	ug/l #	86
55) 1,1,2-Trichloroethane	10.502	97	793	0.54	ug/l #	78
56) Ethyl methacrylate	10.374	69	1228	2.61	ug/l #	90

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031621\
 Data File : VN066212.D
 Acq On : 16 Mar 2021 13:18
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Mar 17 05:21:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
57) 1,3-Dichloropropane	10.644	76	1067	0.41	ug/l #	80
58) 2-Chloroethyl Vinyl ether	9.628	63	55	46.45	ug/l #	54
59) 2-Hexanone	10.695	43	15121	8.38	ug/l	88
60) Dibromochloromethane	10.840	129	371	0.24	ug/l #	42
61) 1,2-Dibromoethane	10.942	107	604	0.42	ug/l	90
64) Tetrachloroethene	10.567	164	1352	0.90	ug/l #	86
65) Chlorobenzene	11.377	112	3090	0.85	ug/l	93
67) Ethyl Benzene	11.454	91	4146	0.65	ug/l #	86
68) m/p-Xylenes	11.564	106	3034	1.28	ug/l	92
69) o-Xylene	11.886	106	1130	0.49	ug/l	89
70) Styrene	11.910	104	2551	2.72	ug/l	98
71) Bromoform	12.069	173	363	0.34	ug/l #	29
73) Isopropylbenzene	12.192	105	4069	0.98	ug/l	99
74) N-amyl acetate	12.020	43	2805	3.33	ug/l #	75
75) 1,1,2,2-Tetrachloroethane	12.447	83	2070	1.43	ug/l #	85
76) 1,2,3-Trichloropropane	12.503	75	2873	1.88	ug/l #	100
77) Bromobenzene	12.471	156	1739	1.60	ug/l	78
78) n-propylbenzene	12.538	91	7343	1.54	ug/l	99
79) 2-Chlorotoluene	12.616	91	3593	1.24	ug/l	74
80) 1,3,5-Trimethylbenzene	12.677	105	4484	1.31	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.345	75	38046	74.67	ug/l #	11
82) 4-Chlorotoluene	12.718	91	4050	1.74	ug/l	95
83) tert-Butylbenzene	12.935	119	3905	1.30	ug/l	97
84) 1,2,4-Trimethylbenzene	12.986	105	5467	1.63	ug/l	94
85) sec-Butylbenzene	13.120	105	7372	1.92	ug/l	100
86) p-Isopropyltoluene	13.235	119	6997	2.08	ug/l	88
87) 1,3-Dichlorobenzene	13.227	146	4355	2.37	ug/l	92
88) 1,4-Dichlorobenzene	13.305	146	5281	2.74	ug/l	85
89) n-Butylbenzene	13.562	91	13856	6.16	ug/l	98
90) Hexachloroethane	13.817	117	621	1.05	ug/l #	38
91) 1,2-Dichlorobenzene	13.592	146	5091	2.81	ug/l	92
92) 1,2-Dibromo-3-Chloropr...	14.214	75	932	3.67	ug/l	66
93) 1,2,4-Trichlorobenzene	14.852	180	14116	11.92	ug/l	98
94) Hexachlorobutadiene	14.949	225	5342	9.28	ug/l	100
95) Naphthalene	15.070	128	108363	31.30	ug/l	100
96) 1,2,3-Trichlorobenzene	15.249	180	17437	14.79	ug/l	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031621\
 Data File : VN066212.D
 Acq On : 16 Mar 2021 13:18
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Mar 17 05:21:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.590 min Scan# 2200
Delta R.T. 0.005 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:168 Resp: 118699
Ion Ratio Lower Upper
168 100
99 51.7 45.4 68.0



#2
Dichlorodifluoromethane
Concen: 0.48 ug/l
RT: 1.821 min Scan# 49
Delta R.T. 0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion: 85 Resp: 566
Ion Ratio Lower Upper
85 100
87 42.0 16.4 49.2









#3
Chloromethane
Concen: 2.39 ug/l
RT: 2.017 min Scan# 122
Delta R.T. -0.002 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 50 Resp: 5056
Ion Ratio Lower Upper
50 100
52 32.0 26.2 39.4



#4
Vinyl Chloride
Concen: 0.72 ug/l
RT: 2.145 min Scan# 170
Delta R.T. -0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion: 62 Resp: 1210
Ion Ratio Lower Upper
62 100
64 35.8 24.9 37.3









#5
Bromomethane
Concen: 2.32 ug/l
RT: 2.529 min Scan# 313
Delta R.T. 0.016 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 94 Resp: 2238
Ion Ratio Lower Upper
94 100
96 72.3 77.3 115.9#



#7
Trichlorofluoromethane
Concen: 0.44 ug/l
RT: 2.969 min Scan# 477
Delta R.T. 0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18


Tgt Ion:101 Resp: 861
Ion Ratio Lower Upper
101 100
103 45.3 50.3 75.5#








#10
Methyl Iodide
Concen: 1.00 ug/l
RT: 3.878 min Scan# 816
Delta R.T. 0.005 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Instrument :
MSVOA_N
ClientSampled :







Tgt Ion:142 Resp: 1486
Ion Ratio Lower Upper
142 100
127 0.0 35.9 53.9#
141 12.0 11.8 17.8



#11
Tert butyl alcohol
Concen: 0.28 ug/l
RT: 4.723 min Scan# 1131
Delta R.T. 0.011 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion: 59 Resp: 83
Ion Ratio Lower Upper
59 100
57 0.0 8.0 12.0#









#13
Acrolein
Concen: Below Cal
RT: 3.532 min Scan# 687
Delta R.T. -0.000 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 56 Resp: 3335
Ion Ratio Lower Upper
56 100
55 97.5 58.0 87.0#



#14
Allyl chloride
Concen: 0.23 ug/l
RT: 4.229 min Scan# 947
Delta R.T. 0.000 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion: 41 Resp: 614
Ion Ratio Lower Upper
41 100
39 98.5 46.4 69.6#
76 66.1 18.8 28.2#





#15

Acrylonitrile

Concen: 3.38 ug/l

RT: 4.886 min Scan# 1192





Delta R.T. 0.002 min

Lab File: VN066212.D

Acq: 16 Mar 2021 13:18

Tgt Ion:	53	Resp:	2957
Ion	Ratio	Lower	Upper
53	100		
52	32.6	66.2	99.2#
51	13.0	28.5	42.7#

Instrument :
MSVOA_N
ClientSampled :



#16

Acetone

Concen: 6.22 ug/l




RT: 3.738 min Scan# 764


Delta R.T. -0.001 min

Lab File: VN066212.D

Acq: 16 Mar 2021 13:18

Tgt Ion:	43	Resp:	5110
Ion	Ratio	Lower	Upper
43	100		
58	27.8	22.3	33.5









#17
Carbon Disulfide
Concen: 3.83 ug/l
RT: 3.969 min Scan# 850
Delta R.T. 0.000 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 76 Resp: 12058
Ion Ratio Lower Upper
76 100
78 8.4 7.3 10.9



#20
Methylene Chloride
Concen: 1.00 ug/l
RT: 4.452 min Scan# 1030
Delta R.T. -0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion: 84 Resp: 1280
Ion Ratio Lower Upper
84 100
49 110.3 130.0 195.0#
51 16.4 39.1 58.7#
86 54.9 50.9 76.3









#21
trans-1,2-Dichloroethene
Concen: 1.55 ug/l
RT: 4.935 min Scan# 1210
Delta R.T. 0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 96 Resp: 1917
Ion Ratio Lower Upper
96 100
61 125.1 133.2 199.8#
98 55.1 52.0 78.0



#25
2-Butanone
Concen: 3.34 ug/l
RT: 6.753 min Scan# 1888
Delta R.T. 0.005 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18


Tgt Ion: 43 Resp: 4268
Ion Ratio Lower Upper
43 100
72 32.2 16.2 24.2#








#27
 cis-1,2-Dichloroethene
 Concen: 0.77 ug/l
 RT: 6.734 min Scan# 1881
 Delta R.T. 0.002 min
 Lab File: VN066212.D
 Acq: 16 Mar 2021 13:18


Instrument :
 MSVOA_N
 ClientSampleId :





Tgt Ion: 96 Resp: 1100
 Ion Ratio Lower Upper
 96 100
 61 90.5 0.0 350.8
 98 32.5 0.0 125.0




#28
 Bromochloromethane
 Concen: 0.41 ug/l
 RT: 7.107 min Scan# 2020
 Delta R.T. -0.000 min
 Lab File: VN066212.D
 Acq: 16 Mar 2021 13:18



Tgt Ion: 49 Resp: 584
 Ion Ratio Lower Upper
 49 100
 129 0.0 0.0 2.4
 130 15.8 43.6 65.4#





#29

Tetrahydrofuran

Concen: 0.95 ug/l

RT: 7.131 min Scan# 2029





Delta R.T. 0.002 min

Lab File: VN066212.D

Acq: 16 Mar 2021 13:18

Tgt Ion:	42	Resp:	835
Ion	Ratio	Lower	Upper
42	100		
72	54.6	26.2	39.4#
71	58.1	25.4	38.2#

Instrument :
MSVOA_N
ClientSampleId :



#30

Chloroform

Concen: 0.38 ug/l




RT: 7.292 min Scan# 2089


Delta R.T. 0.005 min

Lab File: VN066212.D

Acq: 16 Mar 2021 13:18

Tgt Ion:	83	Resp:	936
Ion	Ratio	Lower	Upper
83	100		
85	48.2	49.9	74.9#









#31
Cyclohexane
Concen: 1.25 ug/l
RT: 7.584 min Scan# 2198
Delta R.T. 0.013 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 56 Resp: 3195
Ion Ratio Lower Upper
56 100
69 163.9 23.4 35.0#
84 123.3 57.8 86.6#



#33
1,2-Dichloroethane-d4
Concen: 50.28 ug/l
RT: 7.949 min Scan# 2334
Delta R.T. 0.000 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion: 65 Resp: 85991
Ion Ratio Lower Upper
65 100
67 53.9 0.0 107.2









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.515 min Scan# 2545
Delta R.T. -0.000 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:114 Resp: 202013
Ion Ratio Lower Upper
114 100
63 24.2 0.0 49.2
88 16.3 0.0 33.8



#35
Dibromofluoromethane
Concen: 49.92 ug/l
RT: 7.512 min Scan# 2171
Delta R.T. 0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18


Tgt Ion:113 Resp: 64328
Ion Ratio Lower Upper
113 100
111 102.7 84.0 126.0
192 21.2 16.7 25.1








#36
1,1-Dichloropropene
Concen: 0.78 ug/l
RT: 7.713 min Scan# 2246
Delta R.T. 0.002 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 75 Resp: 1601
Ion Ratio Lower Upper
75 100
110 13.4 16.0 48.0#
77 12.7 24.9 37.3#



#38
Carbon Tetrachloride
Concen: 5.93 ug/l
RT: 7.584 min Scan# 2198
Delta R.T. -0.108 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion: 117 Resp: 10873
Ion Ratio Lower Upper
117 100
119 6.4 74.6 111.8#
121 0.0 23.9 35.9#





#39

Methylcyclohexane

Concen: 0.52 ug/l

RT: 9.008 min Scan# 2729

Delta R.T. -0.003 min

Lab File: VN066212.D

Acq: 16 Mar 2021 13:18





Tgt Ion: 83 Resp: 1152

Ion Ratio Lower Upper

83 100

55 77.7 72.8 109.2

98 23.5 33.7 50.5#



#42

1,2-Dichloroethane

Concen: 0.55 ug/l

RT: 8.051 min Scan# 2372

Delta R.T. 0.002 min

Lab File: VN066212.D




Acq: 16 Mar 2021 13:18


Tgt Ion: 62 Resp: 1223

Ion Ratio Lower Upper

62 100

98 0.0 0.0 14.8









#43
Isopropyl Acetate
Concen: 0.33 ug/l
RT: 8.091 min Scan# 2387
Delta R.T. -0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:	43	Resp:	1473
Ion	Ratio	Lower	Upper
43	100		
61	8.2	14.0	21.0#
87	0.0	7.9	11.9#



#44
Trichloroethene
Concen: 0.53 ug/l
RT: 8.759 min Scan# 2636
Delta R.T. -0.008 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18


Tgt Ion:	130	Resp:	794
Ion	Ratio	Lower	Upper
130	100		
95	71.6	0.0	210.2








#46
Dibromomethane
Concen: 0.34 ug/l
RT: 9.135 min Scan# 2776
Delta R.T. -0.005 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 93 Resp: 337
Ion Ratio Lower Upper
93 100
95 165.9 66.1 99.1#
174 243.3 78.2 117.4#



#47
Bromodichloromethane
Concen: 0.25 ug/l
RT: 9.338 min Scan# 2852
Delta R.T. 0.002 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18


Tgt Ion: 83 Resp: 542
Ion Ratio Lower Upper
83 100
85 44.4 50.3 75.5#
127 0.0 6.2 9.2#








#48
Methyl methacrylate
Concen: 0.65 ug/l
RT: 9.137 min Scan# 2777
Delta R.T. 0.002 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion: 41 Resp: 1392
Ion Ratio Lower Upper
41 100
69 39.1 54.7 82.1#
39 20.8 40.2 60.2#




#49
1,4-Dioxane
Concen: 12.56 ug/l
RT: 9.140 min Scan# 2778
Delta R.T. 0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18



Tgt Ion: 88 Resp: 292
Ion Ratio Lower Upper
88 100
43 0.0 34.4 51.6#
58 0.0 67.3 100.9#









#50
Toluene-d8
Concen: 50.39 ug/l
RT: 10.028 min Scan# 3109
Delta R.T. -0.000 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 98 Resp: 251119
Ion Ratio Lower Upper
98 100
100 63.5 51.0 76.6



#51
4-Methyl-2-Pentanone
Concen: 3.78 ug/l
RT: 9.920 min Scan# 3069
Delta R.T. -0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion: 43 Resp: 10322
Ion Ratio Lower Upper
43 100
58 31.3 27.5 41.3









#52
Toluene
Concen: 0.35 ug/l
RT: 10.089 min Scan# 3132
Delta R.T. -0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 92 Resp: 1250
Ion Ratio Lower Upper
92 100
91 249.7 137.5 206.3#



#53
t-1,3-Dichloropropene
Concen: 0.80 ug/l
RT: 10.320 min Scan# 3218
Delta R.T. -0.000 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18


Tgt Ion: 75 Resp: 1954
Ion Ratio Lower Upper
75 100
77 40.6 25.0 37.4#








#54
cis-1,3-Dichloropropene
Concen: 0.45 ug/l
RT: 9.773 min Scan# 3014
Delta R.T. -0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 75 Resp: 1197
Ion Ratio Lower Upper
75 100
77 15.2 25.0 37.4#
39 61.5 46.2 69.2



#55
1,1,2-Trichloroethane
Concen: 0.54 ug/l
RT: 10.502 min Scan# 3286
Delta R.T. -0.001 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion: 97 Resp: 793
Ion Ratio Lower Upper
97 100
83 67.5 73.3 109.9#
85 34.9 49.8 74.8#
99 61.5 50.9 76.3









#56
Ethyl methacrylate
Concen: 2.61 ug/l
RT: 10.374 min Scan# 3238
Delta R.T. 0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 69 Resp: 1228
Ion Ratio Lower Upper
69 100
41 76.1 66.2 99.4
39 49.3 31.5 47.3#



#57
1,3-Dichloropropane
Concen: 0.41 ug/l
RT: 10.644 min Scan# 3339
Delta R.T. -0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion: 76 Resp: 1067
Ion Ratio Lower Upper
76 100
78 20.4 25.4 38.2#









#58
2-Chloroethyl Vinyl ether
Concen: 46.45 ug/l
RT: 9.628 min Scan# 2960
Delta R.T. -0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 63 Resp: 55
Ion Ratio Lower Upper
63 100
106 0.0 17.6 26.4#



#59
2-Hexanone
Concen: 8.38 ug/l
RT: 10.695 min Scan# 3358
Delta R.T. 0.002 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion: 43 Resp: 15121
Ion Ratio Lower Upper
43 100
58 39.9 23.8 71.5









#60
Dibromochloromethane
Concen: 0.24 ug/l
RT: 10.840 min Scan# 3412
Delta R.T. -0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:129 Resp: 371
Ion Ratio Lower Upper
129 100
127 27.5 39.2 117.6#



#61
1,2-Dibromoethane
Concen: 0.42 ug/l
RT: 10.942 min Scan# 3450
Delta R.T. -0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion:107 Resp: 604
Ion Ratio Lower Upper
107 100
109 104.1 75.5 113.3









#62
4-Bromofluorobenzene
Concen: 41.90 ug/l
RT: 12.345 min Scan# 3973
Delta R.T. 0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 95 Resp: 75257
Ion Ratio Lower Upper
95 100
174 86.0 0.0 162.4
176 77.7 0.0 156.2



#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.347 min Scan# 3601
Delta R.T. 0.000 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion: 117 Resp: 180210
Ion Ratio Lower Upper
117 100
82 59.3 49.0 73.6
119 32.5 25.7 38.5





#64

Tetrachloroethene

Concen: 0.90 ug/l

RT: 10.567 min Scan# 3310

Delta R.T. -0.003 min

Lab File: VN066212.D

Acq: 16 Mar 2021 13:18

Tgt Ion:164 Resp: 1352





Ion Ratio Lower Upper

164 100

166 114.5 98.9 148.3

129 61.5 73.2 109.8#

131 89.6 69.8 104.6



#65

Chlorobenzene

Concen: 0.85 ug/l

RT: 11.377 min Scan# 3612

Delta R.T. 0.003 min

Lab File: VN066212.D




Acq: 16 Mar 2021 13:18


Tgt Ion:112 Resp: 3090

Ion Ratio Lower Upper

112 100

114 35.5 25.4 38.0









#67
Ethyl Benzene
Concen: 0.65 ug/l
RT: 11.454 min Scan# 3641
Delta R.T. 0.002 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 4146
Ion Ratio Lower Upper
91 100
106 37.9 24.2 36.2#



#68
m/p-Xylenes
Concen: 1.28 ug/l
RT: 11.564 min Scan# 3682
Delta R.T. -0.001 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion:106 Resp: 3034
Ion Ratio Lower Upper
106 100
91 223.2 168.2 252.4









#69
o-Xylene
Concen: 0.49 ug/l
RT: 11.886 min Scan# 3802
Delta R.T. -0.006 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:106 Resp: 1130
Ion Ratio Lower Upper
106 100
91 241.2 111.8 335.3



#70
Styrene
Concen: 2.72 ug/l
RT: 11.910 min Scan# 3811
Delta R.T. 0.005 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion:104 Resp: 2551
Ion Ratio Lower Upper
104 100
78 54.9 43.8 65.6
103 57.8 43.7 65.5









#71
Bromoform
Concen: 0.34 ug/l
RT: 12.069 min Scan# 3870
Delta R.T. -0.000 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:173 Resp: 363
Ion Ratio Lower Upper
173 100
175 0.0 23.9 71.8#
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.286 min Scan# 4324
Delta R.T. 0.000 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion:152 Resp: 59042
Ion Ratio Lower Upper
152 100
115 56.9 29.4 88.3
150 154.6 0.0 352.2









#73
Isopropylbenzene
Concen: 0.98 ug/l
RT: 12.192 min Scan# 3916
Delta R.T. -0.000 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:105 Resp: 4069
Ion Ratio Lower Upper
105 100
120 25.1 12.8 38.3



#74
N-ethyl acetate
Concen: 3.33 ug/l
RT: 12.020 min Scan# 3852
Delta R.T. 0.005 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion: 43 Resp: 2805
Ion Ratio Lower Upper
43 100
70 17.2 26.2 39.4#
55 5.3 17.2 25.8#
61 15.4 17.0 25.4#









#75
1,1,2,2-Tetrachloroethane
Concen: 1.43 ug/l
RT: 12.447 min Scan# 4011
Delta R.T. -0.000 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 2070
Ion Ratio Lower Upper
83 100
131 0.0 4.5 13.5#
85 53.4 31.8 95.4



#76
1,2,3-Trichloropropane
Concen: 1.88 ug/l
RT: 12.503 min Scan# 4032
Delta R.T. 0.005 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion: 75 Resp: 2873
Ion Ratio Lower Upper
75 100
77 23.6 0.0 0.0#









#77
Bromobenzene
Concen: 1.60 ug/l
RT: 12.471 min Scan# 4020
Delta R.T. 0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:156 Resp: 1739
Ion Ratio Lower Upper
156 100
77 182.5 115.5 346.4
158 94.8 49.4 148.2



#78
n-propylbenzene
Concen: 1.54 ug/l
RT: 12.538 min Scan# 4045
Delta R.T. 0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion: 91 Resp: 7343
Ion Ratio Lower Upper
91 100
120 21.0 10.8 32.4









#79
2-Chlorotoluene
Concen: 1.24 ug/l
RT: 12.616 min Scan# 4074
Delta R.T. -0.000 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 91 Resp: 3593
Ion Ratio Lower Upper
91 100
126 18.0 16.3 48.8



#80
1,3,5-Trimethylbenzene
Concen: 1.31 ug/l
RT: 12.677 min Scan# 4097
Delta R.T. -0.001 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion: 105 Resp: 4484
Ion Ratio Lower Upper
105 100
120 48.2 23.9 71.8









#81
trans-1,4-Dichloro-2-butene
Concen: 74.67 ug/l
RT: 12.345 min Scan# 3973
Delta R.T. 0.102 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 75 Resp: 38046
Ion Ratio Lower Upper
75 100
53 0.0 77.8 116.6#
89 0.0 33.7 50.5#



#82
4-Chlorotoluene
Concen: 1.74 ug/l
RT: 12.718 min Scan# 4112
Delta R.T. 0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18


Tgt Ion: 91 Resp: 4050
Ion Ratio Lower Upper
91 100
126 28.6 15.7 47.1








#83
tert-Butylbenzene
Concen: 1.30 ug/l
RT: 12.935 min Scan# 4193
Delta R.T. -0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion:119 Resp: 3905
Ion Ratio Lower Upper
119 100
91 64.8 33.7 101.1
134 26.7 12.3 36.9




#84
1,2,4-Trimethylbenzene
Concen: 1.63 ug/l
RT: 12.986 min Scan# 4212
Delta R.T. 0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18



Tgt Ion:105 Resp: 5467
Ion Ratio Lower Upper
105 100
120 47.9 22.1 66.5









#85
 sec-Butylbenzene
 Concen: 1.92 ug/l
 RT: 13.120 min Scan# 4262
 Delta R.T. 0.005 min
 Lab File: VN066212.D
 Acq: 16 Mar 2021 13:18




Instrument :
 MSVOA_N
 ClientSampled :


Tgt Ion:105 Resp: 7372
 Ion Ratio Lower Upper
 105 100
 134 19.1 9.6 28.8



#86
 p-Isopropyltoluene
 Concen: 2.08 ug/l
 RT: 13.235 min Scan# 4305
 Delta R.T. 0.002 min
 Lab File: VN066212.D
 Acq: 16 Mar 2021 13:18

Tgt Ion:119 Resp: 6997
 Ion Ratio Lower Upper
 119 100
 134 16.8 13.1 39.1
 91 26.7 12.0 36.0





#87

1,3-Dichlorobenzene

Concen: 2.37 ug/l

RT: 13.227 min Scan# 4302

Delta R.T. 0.002 min

Lab File: VN066212.D

Acq: 16 Mar 2021 13:18





Tgt Ion:146 Resp: 4355

Ion Ratio Lower Upper

146 100

111 43.3 20.3 61.0

148 73.3 32.4 97.2



#88

1,4-Dichlorobenzene

Concen: 2.74 ug/l

RT: 13.305 min Scan# 4331

Delta R.T. -0.000 min

Lab File: VN066212.D

Acq: 16 Mar 2021 13:18




Tgt Ion:146 Resp: 5281


Ion Ratio Lower Upper

146 100

111 54.1 19.9 59.7

148 58.0 33.2 99.6









#89
n-Butylbenzene
Concen: 6.16 ug/l
RT: 13.562 min Scan# 4427
Delta R.T. 0.005 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91	Resp: 13856
Ion Ratio	Lower Upper
91	100
92	51.3 25.7 77.0
134	27.0 11.8 35.4



#90
Hexachloroethane
Concen: 1.05 ug/l
RT: 13.817 min Scan# 4522
Delta R.T. 0.002 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion: 117	Resp: 621
Ion Ratio	Lower Upper
117	100
201	28.3 42.5 127.5#









#91
1,2-Dichlorobenzene
Concen: 2.81 ug/l
RT: 13.592 min Scan# 4438
Delta R.T. -0.003 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:146 Resp: 5091
Ion Ratio Lower Upper
146 100
111 38.7 20.9 62.7
148 70.4 31.2 93.6



#92
1,2-Dibromo-3-Chloropropane
Concen: 3.67 ug/l
RT: 14.214 min Scan# 4670
Delta R.T. 0.002 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion: 75 Resp: 932
Ion Ratio Lower Upper
75 100
155 54.8 36.5 109.7
157 135.0 46.3 138.8









#93
1,2,4-Trichlorobenzene
Concen: 11.92 ug/l
RT: 14.852 min Scan# 4908
Delta R.T. 0.002 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:180 Resp: 14116
Ion Ratio Lower Upper
180 100
182 93.7 47.4 142.2
145 30.3 14.1 42.3



#94
Hexachlorobutadiene
Concen: 9.28 ug/l
RT: 14.949 min Scan# 4944
Delta R.T. 0.000 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion:225 Resp: 5342
Ion Ratio Lower Upper
225 100
223 62.1 31.3 93.8
227 63.9 31.9 95.7









#95
Naphthalene
Concen: 31.30 ug/l
RT: 15.070 min Scan# 4989
Delta R.T. -0.000 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Instrument :
MSVOA_N
ClientSampled :

Tgt Ion:128 Resp: 108363
Ion Ratio Lower Upper
128 100
127 12.8 10.3 15.5
129 10.6 8.4 12.6



#96
1,2,3-Trichlorobenzene
Concen: 14.79 ug/l
RT: 15.249 min Scan# 5056
Delta R.T. 0.002 min
Lab File: VN066212.D
Acq: 16 Mar 2021 13:18

Tgt Ion:180 Resp: 17437
Ion Ratio Lower Upper
180 100
182 92.8 47.9 143.8
145 30.1 15.6 46.8

