

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031621\
 Data File : VN066220.D
 Acq On : 16 Mar 2021 19:02
 Operator : JC/MD
 Sample : PB135191TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135191TB

Quant Time: Mar 17 05:24:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	263759	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	442934	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	417948	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	176118	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	211763	55.73	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 111.46%			
35) Dibromofluoromethane	7.509	113	143786	50.88	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.76%			
50) Toluene-d8	10.028	98	543467	49.74	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 99.48%			
62) 4-Bromofluorobenzene	12.345	95	178048	45.21	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 90.42%			
Target Compounds					Qvalue	
16) Acetone	3.744	43	90109	49.36	ug/l	97
20) Methylene Chloride	4.452	84	15409	5.40	ug/l	95
25) 2-Butanone	6.756	43	14805	5.22	ug/l	97
43) Isopropyl Acetate	8.094	43	42473	4.28	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031621\
Data File : VN066220.D
Acq On : 16 Mar 2021 19:02
Operator : JC/MD
Sample : PB135191TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135191TB

Quant Time: Mar 17 05:24:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

