

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031621\
 Data File : VN066224.D
 Acq On : 16 Mar 2021 20:34
 Operator : JC/MD
 Sample : M1687-38 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1049-1051

Quant Time: Mar 17 05:25:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.585	168	253673	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	414761	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	384881	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	161967	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	209482	57.32	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	114.64%
35) Dibromofluoromethane	7.507	113	136940	51.75	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.50%
50) Toluene-d8	10.028	98	504683	49.33	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.66%
62) 4-Bromofluorobenzene	12.345	95	162335	44.02	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.04%
Target Compounds						
					Qvalue	
18) Methyl Acetate	4.240	43	11318	2.30	ug/l #	82
37) Ethyl Acetate	6.847	43	58297	9.71	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031621\
Data File : VN066224.D
Acq On : 16 Mar 2021 20:34
Operator : JC/MD
Sample : M1687-38 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1049-1051

Quant Time: Mar 17 05:25:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

