

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031622\
 Data File : VN071311.D
 Acq On : 16 Mar 2022 11:05
 Operator : JC\MD
 Sample : VN0316MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0316MBL01

Quant Time: Mar 16 15:45:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	602261	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1010032	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	903246	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	293857	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	398239	48.247	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.500%
35) Dibromofluoromethane	8.016	113	303074	47.551	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.100%
50) Toluene-d8	10.440	98	1216851	47.418	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.840%
62) 4-Bromofluorobenzene	12.728	95	369456	42.516	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	85.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031622\
Data File : VN071311.D
Acq On : 16 Mar 2022 11:05
Operator : JC\MD
Sample : VN0316MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0316MBL01

Quant Time: Mar 16 15:45:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

