

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031622\
 Data File : VN071318.D
 Acq On : 16 Mar 2022 14:04
 Operator : JC\MD
 Sample : N1905-09ME
 Misc : 5.13g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14A-1A-2-(4.0-4.5)ME

Manual Integrations
 APPROVED

Reviewed By : John Carlone 03/17/2022
 Supervised By : Mahesh Dadoda 03/17/2022

Quant Time: Mar 16 15:49:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.075	168	898011	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.963	114	1447744	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1225454	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	743905	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.428	65	546278	44.386	ug/l	-0.01
Spiked Amount 50.000	Range 61 - 141		Recovery =	88.780%		
35) Dibromofluoromethane	8.016	113	420368	46.013	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	92.020%		
50) Toluene-d8	10.439	98	1988215	54.052	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	108.100%		
62) 4-Bromofluorobenzene	12.727	95	726653	58.339	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	116.680%		
Target Compounds						
31) Cyclohexane	8.075	56	77872m	3.962	ug/l	Qvalue
39) Methylcyclohexane	9.451	83	1050257m	60.992	ug/l	
73) Isopropylbenzene	12.574	105	1223542	17.062	ug/l	98
78) n-propylbenzene	12.916	91	2171582	28.220	ug/l	97
85) sec-Butylbenzene	13.498	105	1269070	19.176	ug/l #	75
89) n-Butylbenzene	13.939	91	1021974	25.106	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031622\
Data File : VN071318.D
Acq On : 16 Mar 2022 14:04
Operator : JC\MD
Sample : N1905-09ME
Misc : 5.13g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14A-1A-2-(4.0-4.5)ME

Quant Time: Mar 16 15:49:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 03/17/2022
Supervised By :Mahesh Dadoda 03/17/2022

