

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031622\
 Data File : VN071325.D
 Acq On : 16 Mar 2022 17:12
 Operator : JC\MD
 Sample : N1886-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2822-B

Quant Time: Mar 17 05:19:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	869849	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1415674	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1354815	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	562732	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	537897	45.120	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	90.240%		
35) Dibromofluoromethane	8.016	113	415821	46.547	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	93.100%		
50) Toluene-d8	10.439	98	1754315	48.774	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	97.540%		
62) 4-Bromofluorobenzene	12.727	95	693178	56.912	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	113.820%		
Target Compounds						Qvalue
11) Tert butyl alcohol	5.363	59	33343	14.237	ug/l #	85
16) Acetone	4.281	43	1655812	285.452	ug/l	96
17) Carbon Disulfide	4.539	76	32402	1.305	ug/l	95
18) Methyl Acetate	4.857	43	239617	15.707	ug/l	97
20) Methylene Chloride	5.098	84	9868	1.306	ug/l	91
25) 2-Butanone	7.333	43	158220	16.932	ug/l	96
29) Tetrahydrofuran	7.686	42	18744	2.896	ug/l #	86
51) 4-Methyl-2-Pentanone	10.327	43	257083	14.063	ug/l	100
52) Toluene	10.504	92	22155	0.834	ug/l	95
59) 2-Hexanone	11.080	43	41220	3.305	ug/l	96
68) m/p-Xylenes	11.945	106	221376	11.763	ug/l	99
95) Naphthalene	15.503	128	18491	5.712	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031622\
Data File : VN071325.D
Acq On : 16 Mar 2022 17:12
Operator : JC\MD
Sample : N1886-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2822-B

Quant Time: Mar 17 05:19:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

