

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031622\
 Data File : VN071330.D
 Acq On : 16 Mar 2022 19:13
 Operator : JC\MD
 Sample : N1957-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MP-3

Quant Time: Mar 17 05:20:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	875979	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1448009	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1358895	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	496020	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	548072	45.651	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.300%
35) Dibromofluoromethane	8.022	113	417526	45.694	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.380%
50) Toluene-d8	10.439	98	1762748	47.914	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.820%
62) 4-Bromofluorobenzene	12.727	95	655892	52.648	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.300%
Target Compounds						
					Qvalue	
18) Methyl Acetate	4.869	43	16968	1.541	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031622\
Data File : VN071330.D
Acq On : 16 Mar 2022 19:13
Operator : JC\MD
Sample : N1957-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MP-3

Quant Time: Mar 17 05:20:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

