

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031622\
 Data File : VN071332.D
 Acq On : 16 Mar 2022 20:01
 Operator : JC\MD
 Sample : N1957-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MP-2

Quant Time: Mar 17 05:21:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	890438	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1462319	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1350635	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	496727	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	554803	45.462	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.920%
35) Dibromofluoromethane	8.022	113	426504	46.220	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.440%
50) Toluene-d8	10.439	98	1772697	47.713	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.420%
62) 4-Bromofluorobenzene	12.727	95	653855	51.971	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.940%
Target Compounds						
					Qvalue	
16) Acetone	4.298	43	9937	1.673	ug/l	96
18) Methyl Acetate	4.851	43	17152	1.535	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031622\
Data File : VN071332.D
Acq On : 16 Mar 2022 20:01
Operator : JC\MD
Sample : N1957-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MP-2

Quant Time: Mar 17 05:21:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

